(4) Basic requirements of the procurement :
A Additional materials for operation table and shadowless lamp must be compatible with existing one.
B Ceiling pendants for the anesthesia apparatus, the endoscope apparatus, the artificial heart lung apparatus have to be provided with matching power supply, LAN, medical gas outlet.
C The surgery video image management system must be able to distribute and record each images taken in the operation room through the network towards other rooms.
D The environmental support management system must be able to maintain operation room environment by contamination control, current monitoring, air condition control etc.
E The anesthesia apparatus must be able to mount a biological information monitoring apparatus and monitor for operation information.
F 2 C-arms are required. One must have 7-inch I.I panel, and the other should have 9-inch I.I panel.
G Neurological function device is required to have more than 8 measurement channel, and it must be able to display both evoked potentials and waveforms on one screen.
H The blood cooler box must be able to manage the optimum temperature of blood.
(5) Time limit for the submission of the requested material : 5 : 00 PM 7, October, 2013
(6) Contact point for the notice : Hideyuki Okihashi, Procurement Section, Accounting Division, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3147

○第６号

1　調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　救急救命システム一式
(3) 調達方法　購入等
(4) 導入予定時期　平成26年度４月以降
(5) 調達に必要とされる基本的な要求要件
A 一次救急から三次救急までの全ての救急患者を受け入れる救急体制がとれる設備であること。
B 患者の生体情報を集中管理出来る設備であること。
C 天吊式の無影灯を有すること。
D 画像診断室へＣＴ撮影装置とＸ線撮影装置の両機を設置し、検査撮影ができる設備であること。
E ＣＴ撮影装置は64列以上のスペックであること。
F 救急医療を行う上で必要な医療器材を効率的に収納・保管ができる棚類を有すること。また、薬品の保存に適した保冷庫を有すること。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　平成25年10月７日17時00分（郵送の場合は必着のこと。）
(2) 提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学病院部総務管理課調達管理係　起橋　秀之　電話0776―61―3111　内線3147
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　平成25年８月21日から平成25年10月７日まで。
(2) 交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1) 開催日時　平成25年８月27日16時30分
(2) 開催場所　福井大学医学部研究棟会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be purchased : Emergency medical system 1 Set
(3) Type of the procurement : Purchase
(4) Basic requirements of the procurement :
A Facilities should be provided that can accept all emergency patients.
B Biological information managing system can be managed collectively.
C Shadowless lump is suspended from the ceiling.
D X-ray CT scanner and X-ray imaging apparatus in the image diagnosis room should be provided for examination.
E Requirements for X-ray CT is 64 slice or higher.
F Storage shelves can store the medical equipment parts effectively. And a refrigerator suitable for preserving the medicine has to be provided.
(5) Time limit for the submission of the requested material : 5 : 00 PM 7, October, 2013
(6) Contact point for the notice : Hideyuki Okihashi, Procurement Section, Accounting Division, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3147

○第７号

1　調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　洗浄・滅菌システム　一式
(3) 調達方法　購入等
(4) 導入予定時期　平成26年度４月以降
(5) 調達に必要とされる基本的な要求要件
A 仕分・洗浄、組立、滅菌保管の３ゾーンが明確に区分けされた設備であること。
B 医療器材の仕分けや組立て作業を行うための準備作業テーブルを有すること。
C エアーガン・ウォーターガンを搭載した流し台を有すること。
D 各洗浄器・滅菌器や流し台にＲＯ水を供給するためのＲＯ水供給機能を有すること。
E 垂直回転棚は医療器材を収納したコンテナおよびセットを収納する事が可能であること。
F 各種機器と連携し、患者への医療器材の使用履歴、洗浄・消毒・滅菌の履歴、滅菌期限、および機器の稼働状況を管理することが可能なトレーサビリティ機能を有した総合滅菌管理システムを含むこと。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　平成25年10月７日17時00分（郵送の場合は必着のこと。）
(2) 提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学病院部総務管理課調達管理係　起橋　秀之　電話0776―61―3111　内線3147
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　平成25年８月21日から平成25年10月７日まで。
(2) 交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1) 開催日時　平成25年８月29日13時00分
(2) 開催場所　福井大学医学部研究棟会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be purchased : Disinfection and sterilization system 1 Set
(3) Type of the procurement : Purchase
(4) Basic requirements of the procurement :
A The facility area has to be provided with clearly separated 3 zones : Sorting／cleaning zone for soiled items, Packing zone for cleaned items, and Sterile Stock.
B Preparation Tables should be provided for sorting and packing of medical instrument／equipment.
C The sink unit equipped with a spray air gun and a spray water gun has to be provided.
D RO water production system has to be provided, which can supply RO water to each washer disinfector, sterilizer, and sink unit.
E The Vertical carousel has to be provided, which should have space to keep sterile containers for medical instrument.