② This system should provide the communication tools and data format which are based on the international standard measure, and will exchange the patient basic information, medical care records and image information smoothly.

③ This system should have a function to stock and manage patient clinical data and information (which are independent from data base for business use) in the format based on the international standard measures for long term.

④ This system should possess enough data processors, terminals and terminal and networks to archive a required function.

⑤ This system should have interfaces to connect with the equipment and systems already used in each section of the hospital.

⑥ The part of this system which is indispensable for the hospital business must be operated automatically for 24 hours a day, even if it is unmanned, and always operated safely and easily, Also it should take enough measures to coop with the system failure.

⑦ This system should have a function to restrict access based on users. also it should take enough measures to secure privacy and protect security.

⑧ This system must be based on the latest technology to support hospital business and medical services. And the operability and response time must be appropriate for the actual medical care. In addition the secondary utilization of each data must be possible.

⑨ For a stable of this system, 24 hour a day monitoring and quick response against failure are needed.

⑸ Time-limit for the submission of the requested material : 3 : 00 PM 3, March 2014

⑹ Contact point for the notice : Hideyuki Takami Procurement Section, Accounting Division, Kushiro Rosai Hospital 13―23 Nakazono-cho Kushiro-shi Hokkaido 085―8533 Japan TEL 0154―22―7191

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成26年1月31日

独立行政法人国立病院機構

本部総務部長　佐藤　美幸

◎調達機関番号　597　◎所在地番号　13

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量

A　マルチスライスＣＴ　6式

B　ＭＲＩ　6式

C　血管連続撮影装置　7式

D　ガンマカメラ　3式

E　Ｘ線透視撮影装置　6式

F　Ｘ線一般撮影装置　6式

G　乳房Ｘ線撮影装置　2式

H　外科用イメージャー　3式

I　高度放射線治療装置　2式

⑶　調達方法　購入（国立病院機構および労働者健康福祉機構における共同入札）

⑷　導入予定時期

A～H　平成26年度第2・四半期以降

I　平成27年度

⑸　調達に必要とされる基本的な要求要件

A　マルチスライスＣＴ　6式

①　1スキャンで64スライス以上撮影することができること。

②　全肺や全腹部を数秒で撮影する能力があること。

B　ＭＲＩ　6式

①　1.5テスラ以上の磁場強度を有する磁気共鳴画像装置であること。

②　頭部、躯幹部、四肢などすべての検査で高精細な画像が迅速に撮影でき、解析機能が充実していること。

③　高速処理可能な高性能な操作コンソールを有し、操作性に優れていること。解析診断用のワークステーションを備えていること。

C　血管連続撮影装置　7式

①　不整脈の診断、治療と頭腹部及び下肢の血管造影による診断、治療および全身の血管造影による診断、治療が安全に行えること。

②　Ｃアームは1方向又は2方向を有し、迅速にアームの角度を設定でき、任意の方向からデジタル透視、デジタル撮影ができること。

③　フラットパネルディテクタは、血管造影による診断・治療が行えること。

④　デジタル画像処理システムは血管内治療を支援する機能や解析機能を有すること。

⑤　患者、術者、スタッフに対する被ばく低減機構を有すること。

D　ガンマカメラ　3式

①　多検出器デジタルガンマカメラ装置であること。

②　スペクト収集とホールボディ収集の機能を有していること。

③　フュージョンソフトや定量性を行えるアプリケーションを十分に備えていること。

E　Ｘ線透視撮影装置　6式

①　透視撮影ができること。

②　動画記録が可能であること。

③　寝台の水平位が可能であること。

④　フラットパネルディテクタを有すること。

F　Ｘ線一般撮影装置　6式

①　Ｘ線管球を有すること。

②　立位、臥位の撮影台を有し、全身の撮影が迅速かつ安全な撮影への対応が可能であること。

③　フラットパネルディテクタを有すること。

G　乳房Ｘ線撮影装置　2式

①　デジタル乳房Ｘ線撮影装置であること。

②　乳房バイオプシができるシステムを有すること。

③　乳房撮影診断用高精細モニタシステムを有すること。

H　外科用Ｘ線撮影装置　3式

デジタル外科用Ｘ線撮影装置であること。

I　高度放射線治療装置　2式

①　多目的な治療に対応が可能であること。

②　遮蔽能力の高い多分割コリメータを有すること。

③　システムとして安全で簡便な操作性を有し、高精度な放射線治療への対応が可能であること。

2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成26年3月17日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒152―8621 東京都目黒区東が丘2―5―21　独立行政法人国立病院機構本部財務部整備課長　河田　晃伸　電話03―5712―5071

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成26年1月31日から平成26年3月17日まで。

⑵　交付場所　上記2⑵に同じ。

4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased :

A　Multi-slice Computed Tomography System 6 sets

B　Magnetic Resonance Imaging System 6 sets

C　Serial Angiographic System 7 sets

D　Gamma Camera System 3 sets

E　X-ray fluoroscope Radiography System 6 sets

F　General Radiographic System 6 sets

G　Digital X rays equipment to mammography 2 sets

H　Digital Mobile Imaiging System 3 sets

I　High Radiationtherapy System 2 sets