⑶ Type of the procurement : Purchase
⑷ Basic requirements of the procurement :
A Multi-slice Computed Tomography System :
① The system should be capable of scanning more than 64 slices data during one rotation.
② The system should have ability that could scan either whole lung or abdomen in few seconds.
B Magnetic Resonance Imaging System :
① Magnetic field strength should be 1.5 Tesla or more.
② The images of examination in head, body and extremities are enable to examine speedily with high resolution images.
③ The console unit must provide high performance capability and exceptional in operation.
C Serial Angiographic System :
① The system should be designed safely for interventional procedures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography.
② C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary-direction.
③ The flat-panel detector should be especially designed to fulfill the interventional angiography.
④ The digital imaging processing system should have supporting and analysis function for intravascular treatment.
⑤ The system should equip the radiation exposure reduction unit for patient, operator, and staff.
D Gamma Camera System :
① The system should be multiple digital gamma camera.
② The system should have function of SPECT and Whole body acquisition.
③ The system should equip the fusion software and quantitative application-replete.
E X-ray fluoroscope Radiography System :
① The system is available in fluoroscopic radiography.
② The system should capable of recording dynamic images.
③ The system should be capable of positioning the bed position to horizontal.
④ The system should equip the flat-panel detector.
F General Radiographic System :
① The system should equip x-ray tube.
② The system should equip a table forstanding and lying position, in addition, radiography of whole body couldable to perform speedily and safely.
③ The system should equip the flat-panel detector.
G Digital X rays equipment to mammography :
① The system should be digital X rays equipment to mammography.
② The system must be able to digital stereotactic mammography.
③ The system should equip the high resolution monitor system for mammography diagnosis.
H Digital Mobile Imaiging System :
The system should be digital mobileimaiging system.
I High Radiationtherapy System :
① The system should be capable to multipurpose treatment.
② The system should equip multi-leaf collimator with high shielding performance.
③ The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiationtherapy.
⑸ Time-limit for the submission of the requested material : 5 : 00 P.M. March 17, 2014
⑹ Contact point for the notice : Terunobu Kawata, Manager, Facilities Planning Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2―5―21 Higashigaoka Meguro-ku Tokyo-to 152―8621, Japan. TEL 03―5712―5071

## 意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成26年1月31日
総務省大臣官房会計課企画官　梅田　　勉
◎調達機関番号　012　◎所在地番号　13
1　調達内容
⑴　品目分類番号　14
⑵　借入物品等及び数量　電子政府の総合窓口(e-Gov)システムに係る機器及びソフトウェアの借入　1式
2　意見の提出方法
⑴　意見の提出期限　平成26年2月21日午後5時まで
⑵　提出先　〒100―8926 東京都千代田区霞が関2－1－2　総務省大臣官房会計課契約第一係　宮下　浩平　電話03―5253―5132
3　仕様書案の交付等
⑴　交付期間　本日から平成26年2月21日まで
⑵　交付場所　総務省ＨＰの調達情報から総務省電子入札システムへアクセスし入手すること。
4　Summary
⑴　Classification of the products to be procured : 14
⑵　Nature and quantity of the products to be leased : The leasing equipment for the "e-Gov" Information Portal system and electronic application system 1 set.
⑶　Time limit for the submission of comments : 5 : 00 PM 21 February 2014.
⑷　Contact point for the notice : Kohei Miyashita, Accounts Division, Minister's Secretariat, Ministry of Internal Affairs and Communications, 1―2 Kasumigaseki 2-chome Chiyoda-ku Tokyo 100―8926 Japan TEL 03―5253―5132.

## 意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成26年1月31日
日本郵便株式会社調達部担当執行役員
河村　　学
◎調達機関番号　411　◎所在地番号　13
**○第１号（№１）**
1　調達内容　⑴品目分類番号　9　⑵購入等物品及び数量　ロールパレット(鉄製・Ⅱ型(改))1,000台（予定）
2　意見の提出方法　⑴意見の提供期限　平成26年2月21日午前11時00分（郵送の場合は必着のこと）⑵提出先　〒100―8798 東京都千代田区霞が関一丁目3番2号　日本郵便株式会社調達部(物品契約担当)担当　津田　紀彦　電話03―3504―9532
3　仕様書案の交付　⑴交付期間　平成26年1月31日から平成26年2月20日午後3時00分まで　⑵交付場所　上記2⑵に同じ
4　Summary　⑴Classification of the services to be procured : 9　⑵Nature and quantity of the products to be purchased : About 1,000 pieces of Roll-box-pallets (Steel Type 2 Advanced type)　⑶Time limit for the submission of comments : 11 : 00 AM 21 February 2014　⑷Contact point for the notice : Norihiko Tsuda, Procurement Department, Japan Post Co., Ltd. 1―3―2 Kasumigaseki Chiyoda-ku, Tokyo 100―8798 Japan TEL 03―3504―9532
**○第２号（№２）**
1　調達内容　⑴品目分類番号　9　⑵購入等物品及び数量　ロールパレット(鉄製・Ⅴ型(改))45,100台（予定）
2及び3　第1号の2及び3に同じ
4　Summary　⑴, ⑶ and ⑷The same with the notice above (No. 1)　⑵Nature and quantity of the products to be purchased : About 45,100 pieces of Roll-box-pallets (Steel Type 5 Advanced type)