第明三治十種二五郵年便三物月認三可十一日

⑸　調達に必要とされる基本的な要求要件
①　検出器は対面型で２個装備し、現行装置以上の性能を有する機器であること。
②　散乱線補正を有し、高分解能で高画質であること。
③　画像の重ね合わせに優れ、効率的な解析が行えること。
④　高次元のソフトウェアを有し、機能拡充に対応できること。
⑤　既存のＨＩＳ・ＲＩＳおよびＰＡＣＳに接続できること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成26年４月７日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒860―0008 熊本県熊本市中央区二の丸１－５　独立行政法人国立病院機構熊本医療センター　企画課長　柳田　和憲　電話096―353―6501　内線2320

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成26年２月21日から平成26年４月７日まで。

⑵　交付場所　上記２⑵に同じ。

4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Gamma camera 1 set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :
①　The detector should be apparatus which equips two pieces with a met type and has the performance more than the present equipment.
②　It has scattered radiation compensation and is high resolution and high definition.
③　Excel in superposition of a picture and conduct efficient analysis.
④　The system have a high-dimensional software, cope with function expansion.
⑤　The system must be connected to the hospital information systems, RIS (Radiology Information System) and PACS (Picture Archiving and Communication System).

⑸　Time-limit for the submission of the requested material : 5 : 00 P.M. April 7, 2014

⑹　Contact point for the notice : Kazunori Yanagida, Director, Accounting Division, National Hospital Organization Kumamoto Medical Center, 1―5 Ninomaru Chuo-ku Kumamoto-shi Kumamoto-ken 860―0008, Japan. TEL 096―353―6501 ext. 2320

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 26 年２月 21 日

国立大学法人旭川医科大学長　吉田　晃敏

◎調達機関番号　415　◎所在地番号　01

○第１号

1　調達内容

⑴　品目分類番号　14

⑵　導入計画物品及び数量　情報基盤センター電子計算機システム　一式

⑶　調達方法　借入

⑷　導入予定時期　平成26年度２月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　情報基盤センター電子計算機システムは、医学・看護学研究支援サブシステム、情報処理教育支援サブシステム、学術情報支援サブシステム、管理運用サブシステムの４サブシステムにより構成し、本学の研究・教育活動の強化・高度化を支援するシステムであること。

Ｂ　医学・看護学研究支援サブシステムは、医学・看護学研究におけるデータの共有並びに総合的解析を進めるための数値計算、画像解析、生体信号解析等、最先端の研究活動を支援するシステムであること。

Ｃ　情報処理教育支援サブシステムは、全学生の登録管理の機能を装備し、すべての端末計算機を同時に使用した場合でも即応的な応答が得られるシステムで、かつ情報リテラシー教育、医療情報学教育等、マルチメディアに対応した情報処理教育の実施が可能なシステムであること。

Ｄ　学術情報支援サブシステムは、電子図書館機能の充実・強化を支援するシステムであること。

Ｅ　管理運用サブシステムは、他のサブシステムと有機的に相互接続し、システムの一元的管理運用を行うシステムであること。

Ｆ　すべての計算機は最新の各種技術を用い十分高速なネットワークを介して接続されること。

Ｇ　システムは、既存のキャンパス情報ネットワークと接続し、コンピュータ資源の共有化が図られること。

Ｈ　計算機は、十分なセキュリティ機能を有し、利用者層に応じたシステム資源の利用権の設定が可能なこと。

Ｉ　システムの障害を未然に防止し、障害発生に対しては即時的に対処でき、安定した運用が可能なこと。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成26年３月25日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒078―8510 北海道旭川市緑が丘東２条１丁目１－１　旭川医科大学総務部会計課調達係長　吉田　誠一　電話0166―68―2157

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成26年２月21日から平成26年３月25日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成26年３月４日14時00分

⑵　開催場所　旭川医科大学事務局第二会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 14

⑵　Nature and quantity of the products to be rent : The Information and Communication Technology Center Computer System 1 Set

⑶　Type of the procurement : Rent

⑷　Basic requirements of the procurement :

A　The Information and Communication Technology Center Computer System should be composed of four systems. i.e., a medical and nursing research subsystem, an information processing education subsystem, a library information subsystem, and a network administration subsystem. It should support advanced research and education.

B　The medical and nursing research subsystem should support advanced research such as computing, image analysis, biological signal analysis for sharing data and integrated analysis in medical and nursing research.

C　The teaching of information processing education subsystem should have a function that can register and administer all undergraduates and provide simultaneous use of all the terminals. It should be able to provide a multimedia education such as computer literacy and medical information.

D　The library information subsystem should support reinforcement and adequate ability of a digital library.

E　The administration subsystem should connect the other subsystems systematically and provide the capability to unify the administration of all computer and network systems.

F　All Computers should be connected to one another by a sufficiently fast network with the latest technology.

G　The system should be connected to existing campus information network and all computer resource should be shared.

H　Computers should support high security and resource allocation facilities according to the user level.

I　The system should include a mechanism to prevent the occurrence of system failure, and provide the capability for stable operation and instantaneous recovery from a system failure.