○第３号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　血液浄化総合システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成26年度10月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　逆浸透精製水製造システム

ａ　原水タンク、原水ポンプ、プレフィルタ、軟水機、カーボンフィルタ、ＲＯモジュール、ＲＯポンプ、ＲＯ水タンク及び送水ポンプにて構成されていること。

Ｂ　全自動溶解装置

ａ　Ａ剤及びＢ剤の粉末型透析用剤を自動溶解し、透析原液を供給できること。

Ｃ　個人用透析装置

ａ　一人の患者の血液透析を行うために透析用監視装置の機能と透析液調整機能を備えた装置であること。

Ｄ　個人用多用途透析装置

ａ　一人の患者の血液透析を行うために透析用監視装置の機能と透析液調整機能を備えた装置であること。

ｂ　血液透析、血液透析濾過、血液濾過等を行うことができる個人用透析装置であること。

Ｅ　多用途透析用監視装置

ａ　血液透析器を用いた慢性又は急性腎不全等腎機能が著しく低下した症例の血液透析（ＨＤ）治療において、透析液（温度・流速等）の監視、除水及び血液体外循環の制御及び監視を行うことができる装置であること。

Ｆ　血液浄化装置（Ａ）

ａ　持続緩徐式血液濾過透析（ＣＨＤＦ）、持続緩徐式血液濾過（ＣＨＦ）、持続緩徐式血液透析（ＣＨＤ）を主機能としていること。

Ｇ　血液浄化装置（Ｂ）

ａ　血液浄化療法適応患者に対し、安全且つ簡便な血液浄化を行うことができること。

ｂ　治療モードとして、持続緩徐式血液濾過（ＣＨＤＦ・ＣＨＤ・ＣＨＦ・ＳＣＵＦ）、アフェレーシス（ＰＥ・ＤＦＰＰ・ＰＡ・ＨＡ・ＬＣＡＰ）、腹水濾過が選択できること。

Ｈ　微量血液凝固計

ａ　抗凝固療法を行う際にモニタリングとして使用される凝固時間（ＡＣＴ）が測定可能な装置であること。

Ｉ　超音波診断装置

ａ　超音波診断装置は、血管内や体内を超音波により描出する器械で、透析中でも効率よく検査が行なえるように持ち運びが簡便なポータブルタイプであること。

Ｊ　セントラルモニタ

ａ　セントラルモニタは、本学の東西病棟で使用中の生体情報モニタ又は、送信機を併用で監視可能なモニタであること。

Ｋ　生体情報モニタ

ａ　生体情報モニタは、本学の東西病棟で使用中のセントラルモニタで送信可能なモニタであること。

Ｌ　体成分分析装置

ａ　疾患者に重要な体水分や筋肉等の体成分情報を部位別に高い精度で分析できること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成26年５月27日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒860―8556　熊本市中央区本荘１－１－１　熊本大学医学部附属病院事務部財務・調達ユニット　上野　孝之　電話096―373―5938

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成26年４月11日から平成26年５月27日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成26年４月22日14時30分

⑵　開催場所　熊本大学医学部附属病院第二会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Blood purification integrated system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Reverse osmosis purified water manufacturing system

a　The system must be consisted by Raw water tank, Raw water pump, Pre-filter, Soft water machine, RO module, RO pump, RO water tank and Water tank.

B　Full automatic dissolution equipment

a　Can automatically solve the powder form of dialysis for agent B and agent A, and it is possible to supply the dialysis solution.

C　Personal dialysis machine

a　The unit is equipped with the dialysate adjustment features and functions of dialysis monitoring device in order to perform the hemodialysis of one patient.

D　Personal versatile dialysis machine

a　The unit is equipped with the dialysate adjustment features and functions of dialysis monitoring device in order to perform the hemodialysis of one person.

b　It is a personal dialysis device which can perform hemodialysis, hemodiafiltration and hemofiltration.

E　Versatile dialysis monitoring device

a　Hemodialysis patients with acute renal failure, such as renal function or chronic using blood dialyzer was significantly reduced in (HD) treatment, it is a device that can do monitoring of dialysis (temperature, flow speed etc), monitoring and control of blood extracorporeal circulation and water removal.

F　Blood purification equipment (A)

a　They are the main feature persistent slow-type hemodiafiltration (CHDF), sustained expression slow hemofiltration (CHF), sustained expression slow hemodialysis (CHD).

G　Blood purification equipment (B)

a　It performs a simple and safe blood purification to the blood purification therapy adaptive patient.

b　As a treatment mode, is sustains continuous slow type blood filtration (CHDF, CHD, CHF, SCUF), Apheresis (PE, DFPP, PA, HA, LCAP), ascites filtration can beselected.

H　Trace blood coagulation meter

a　It is capable to measure clotting time (ACT) that is used for monitoring when performing anticoagulation therapy.

I　Ultrasonic diagnostic equipment

a　Ultrasound diagnostic device equipment, in the instrument to render ultrasonically body and blood vessel, that is relatively simple to carry portable type to perform so efficiently even during dialysis.

J　Central monitor

a　Central monitor needs to have the ability to monitor the patient monitoring system or transmitter currently being used in the East and West hospital ward in the Kumamoto University Hospital.

K　Biological information monitor

a　Biological information monitor is required to connect to the Central Monitor used in the East and West hospital ward in the Kumamoto University Hospital.

L　Body composition analyzer

a　It can be analyzed with high precision separate sites the body component of important information such as muscle and body water to the disease's.

⑸　Time limit for the submission of the requested material : 17 : 00　27 May, 2014

⑹　Contact point for the notice : Takayuki Ueno, Finance and Procurement Division, Hospital Administration Office, Kumamoto University, 1―1―1 Honjo, Chuo-ku, Kumamoto-shi 860―8556 Japan, TEL 096―373―5938