4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成26年７月１日14時30分
⑵　開催場所　香川大学医学部管理棟４階会議室１
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : High Pressure Steam Sterilizer 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　High Pressure Steam Sterilizer
a　Have a built-in steam generator for high quality steam and the generated high quality steam there should be able to be used its sterilization process.
b　Have a door each at entry and exit side for passing through a load.
B　Floor Loading Cart
a　Designed particularly for the steam sterilizer. Shall contain more than three shelves, and be loaded directly into the sterilizer.
C　Steam Sterilizer Management System
a　Shall manage the sterilizer operating data on a real-time basis, enables to record the operating history, input the pass-fail result of the sterilization, print out the data sheet, and indicate the operating error.
⑸　Time limit for the submission of the requested material : 17 : 00　8 August, 2014
⑹　Contact point for the notice : Hiroyuki Nitta, Procurement Section, Accounting Division, Kagawa University, 1750―1 Oaza Ikenobe Miki-cho Kita-gun Kagawa 761―0793 Japan, TEL 087―891―2492

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 26 年６月 23 日
国立大学法人香川大学長　長尾　省吾
◎調達機関番号　415　◎所在地番号　37
○第４号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　手術室映像管理システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成27年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　手術室映像管理システム
a　術野カメラは、無影灯灯体もしくは無影灯カメラアームに取り付けられ、ハイビジョン映像対応であること。
b　手術室内のハイビジョンディスプレイに、内視鏡カメラや手術顕微鏡カメラなどの術野映像、患者生体情報、画像ネットワーク（ＰＡＣＳ）、病院情報管理システム（ＨＩＳ）を切り換えて表示できること。
c　ハイビジョンディスプレイは、天井懸架式及び壁面埋込式であること。
d　タッチパネル等によってカメラ制御・画像配信蓄積などの操作を容易に行えること。
e　術野映像は、デジタル画像化しリアルタイムに映像配信できること。
f　術野映像のデジタル画像を保存・蓄積できること。
g　画像を編集できること。
h　病院情報管理システムと接続し、患者基本情報を取得できること。
i　画像を病院情報管理システム端末へ配信できること。
B　無影灯
a　天井懸垂式であること。
b　光源はＬＥＤを使用すること。
c　無影灯用アームは360度回転可能なこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成26年８月８日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒761―0793 香川県木田郡三木町大字池戸1750―１　香川大学医学部管理課専門職員（契約）　新田　浩幸　電話087―891―2492
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成26年６月23日から平成26年８月８日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成26年７月１日14時00分
⑵　開催場所　香川大学医学部管理棟４階会議室１
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Image management system for operating room 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Image management system for operating room
a　The operative field camera should be a high definition video camera that can be mounted to shadowless lamp body or shadowless lamp camara arm in the operating room.
b　The high difiniton display in the operating room can be utilized as the terminal display for essential information such as endoscope image, operating microscope image, patient monitor, picture archiving and communication system (PACS), hospital information system (HIS).
c　The high difiniton display should be ceiling suspension type and wall embedding type.
d　Operation of the system should be done easily by touch panel, etc.
e　Image of surgical site should be digitalized and provided in real time.
f　Digital images of surgial site should be able to saved and　stored.
g　The system should have feature of processing image.
h　The system should be connected with hospital information system (HIS) and be capable of acquiring the patient basic information.
i　The system should have a feature of distributing the data to the hospital information terminal.
B　Shadowless lamp
a　is a ceiling chinning-exercises type.
b　A light source is using LED.
c　The arm for shadowless lamps is that 360-degree rotation is possible.
⑸　Time limit for the submission of the requested material : 17 : 00　8 August, 2014
⑹　Contact point for the notice : Hiroyuki Nitta, Procurement Section, Accounting Division, Kagawa University, 1750―1 Oaza Ikenobe Miki-cho Kita-gun Kagawa 761―0793 Japan, TEL 087―891―2492