H　The system must provide enriched functions compared with our current system. More concretely, the system must succeed all clinical data of the current system and functions which are necessary for our clinical processes, and the system must enable users to access all clinical data on the current system.

I　The system must equip sufficient terminals including some necessary high-resolution monitors. The system must provide sufficient devices for support of our clinical processes. Moreover, the system must equip functions managing these hardware entirely.

J　The system must equip functions allowing the clinical process PDCA (plan-do-check-act) cycle to work quickly and flexibly. More concretely, the system must equip functions updating the system itself periodically through discussion with our staff.

K　The system must be secure. More concretely, the system must ensure the security of clinical records and personal privacy using VDI (virtual desktop infrastructure) technology. Moreover, the system must equip functions providing clinical data generated in our hospital to other clinical institutions within a secured environment, and functions managing access records of the data.

L　The system must include system support. More concretely, the system must include a maintenance system for the whole system and avoid system failures. In case a system failure happens, the system must be restored promptly. Additionally, the system must provide easy-to-understand manuals and sufficient operation training, including training to enable our staff to operate the system non-stop.

⑸　Time limit for the submission of the requested material : 17 : 00　7　November, 2014

⑹　Contact point for the notice : Mitsuhiro Fukuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 26 年 10 月 1 日

国立大学法人高知大学長　脇口　　宏

◎調達機関番号　415　◎所在地番号　39

○第１号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　検体系検査統合システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成27年度１月以降

⑸　調達に必要とされる基本的な要求要件

A　検体検査搬送システムは、検体回収機、検体投入機、自動開栓機、自動分注機、自動閉栓機、多項目血球分析装置、血液塗末標本作成装置、血液凝固分析装置、免疫分析装置、生化学分析装置、検体収納冷蔵庫を備え、連動して測定ができること。また、障害時の対応として分析装置を切り離して使用できること。

B　臨床検査情報システムは、サーバー、デスクトップ型パソコン、ノート型パソコン、レーザプリンタ、ハンディバーコードリーダー、ＩＤカードリーダー、ソフトウェア一式を備え、既存病院システムと接続し稼動できること。

C　採血業務システムは、採血管準備装置、採血台、採血整理番号呼び込み表示装置を備えること。採血管準備システムは既存病院情報システムと接続し稼動できること。

D　付帯設備工事は外部からの侵入に対してセキュリティーが確保できるものとし、検査部内部は検体・人員等の動線が確保できる構造であること。また、今回、導入するシステムが稼動できる必要な電源等の工事費用は含んでいること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成26年11月17日17時15分（郵送の場合は必着のこと。）

⑵　提供先　〒783―8505 南国市岡豊町小蓮　高知大学医学部・病院事務部会計課専門職員　中山　浩行　電話088―880―2241

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成26年10月１日から平成26年11月17日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成26年10月15日15時00分

⑵　開催場所　高知大学医学部・病院事務部会計課会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : New generation total laboratory management system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The Laboratory Automation system must be equipped with unloading and sorting module, sample loading module, decap connector module, aliquot connector module, re-cap connector module, automated hematology analyzer, fully automated slide/maker analyzer, automated coagulation analyzer, immunological auto analyzer, biochemical auto analyzer, and sample storage refrigerator. All of these analyzers must be linked together to measure the samples. Moreover, in case of trouble, each analyzer has to used in isolation.

B　Clinical laboratory information system must have a server, desktop PC, notebook sized PC, laser beam printer, bar code label printer, handy code reader, ID card reader, and a set of software which is connected and work with the existing hospital information system.

C　The blood collecting operation system must be equipped with a blood collecting pipe preparation equipment, blood collecting stand and a blood collecting reference number calling-in display system. The blood collecting operation system must be connected and work with the existing hospital information system.

D　Incidental-facility construction must be secured from outside invasion. The clinical laboratory should be structured to secure the flow of the samples and the staff. Moreover, other construction expenses to work this system, such as power supply, must be included.

⑸　Time limit for the submission of the requested material : 17 : 15　17　November, 2014

⑹　Contact point for the notice : Hiroyuki Nakayama, Expert Official, Accounting Division, Kochi Medical School, Kochi University, Kohasu Oko-cho Nankoku-shi 783―8505 Japan, TEL 088―880―2241

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 26 年 10 月 1 日

国立大学法人九州大学総長　久保　千春

◎調達機関番号　415　◎所在地番号　40

○第４号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　シリンジポンプ　300台

⑶　調達方法　借入

⑷　導入予定時期　平成27年度７月以降

⑸　調達に必要とされる基本的な要求要件

⑷駆動装置がスライダー等を介してシリンジのピストンを押し込み、持続的に設定した流量で薬剤等を注入する装置であること。⑻点滴スタンド等に装着して使用する機能を有すること。⑹誤作動を防止する安全機能や、異常時のアラーム機能を有すること。