6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Magnetic Resonance Imaging and X-ray Computed Tomography System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　Computed Tomography System
a　This system should have multi-slice scanning facilities with date acquisition system of more than 64 slices per rotation.
b　This system should have a capability to obtain images with extremely high temporal and spatial resolutions.
c　This system should have a capability of rapid acquisition of volume data covering long distance of body axis.
d　This system should have a excellent capability of morphologic and functional study for cardiovascular diseases.
e　This system should be carefully designed on patient safety and amenity, as well as space utility.
f　This system should have easy operability, rapid image processing facilities, and user-friendly diagnostic screen function.
B　Magnetic Resonance Imaging System
a　This system should be a high magnetic filed (3tesla) magnetic resonance imaging system that can provide images of whole human body with High-resolution quality.
b　This system must allow for multi-coil function and rapid acquisition of high resolution images with parallel imaging technique.
c　This system must allows for acquisition of not only anatomical imagingbut also functional and metabolic information of all human organs, including perfusion imaging, functinal imaging and MR spectroscopy (including MR spectroscopic imaging)
d　This system should have an user-friendly interface, allowing for the fast and easy processing of analysis and display of all kinds of date sets.
e　This system have capability to store and communicate images and patient's informaition as standard image formats including DICOM.
⑸　Time limit for the submission of the requested material : 17 : 00　6 February, 2015
⑹　Contact point for the notice : Norioki Ito, Procurement Section, Niigata University Medical and Dental Hospital, 754 Asahimachidori 1-bancho Chuo-ku Niigata-shi 951―8520 Japan, TEL 025―227―0970

○第４号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　全身用コンピュータ断層撮影装置　一式
⑶　調達方法　借入
⑷　導入予定時期　平成27年度10月以降
⑸　調達に必要とされる基本的な要求要件
全身用コンピュータ断層撮影装置
A　64列以上の検出器を有すること。
B　既存の画像サーバー、放射線部門情報管理システムと連動すること。
C　各種形態解析、機能解析が行えること。
D　全身の等方性のボリュームデータを撮影可能であること。
E　ガントリ開口部は十分な開口経を有していること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提出期限　平成27年２月６日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課契約係　伊藤　憲興　電話025―227―0970
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成26年12月22日から平成27年２月６日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成27年１月５日14時00分
⑵　開催場所　新潟大学医歯学総合病院西診療棟３階　第２会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Computed Tomography Scanner System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement : Computed Tomography Scanner System
A　64 or more multi-slice capability.
B　Connectable with existing network system.
C　Various morphological and functional analyzing software.
D　The system should be capable of is otropic volume CTscan of the whole body.
E　The bore size of CT gantry shall be larger than normal type.
⑸　Time limit for the submission of the requested material : 17 : 00　6 February, 2015
⑹　Contact point for the notice : Norioki Ito, Procurement Section, Niigata University Medical and Dental Hospital, 754 Asahimachidori 1-bancho Chuo-ku Niigata-shi 951―8520 Japan, TEL 025―227―0970

○第５号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　核医学画像診断システム　二式
⑶　調達方法　借入
⑷　導入予定時期　平成27年度10月以降
⑸　調達に必要とされる基本的な要求要件
A　デジタルガンマカメラ【A】
a　本システムは二検出器ガンマカメラを備えた核医学画像診断システムであり、ＳＰＥＣＴ断層撮影により様々な放射性医薬品を利用できること。
b　高画質・高速撮影の出来るＸ線ＣＴ装置が組み合わされたＳＰＥＣＴ装置を有すること。
c　脳、心臓領域、及び全身検査に対応可能であること。
d　操作性に優れ、画像処理及び機能解析が高速かつ容易に行えること。
e　画像データ及び関連情報の転送保管がＤＩＣＯＭ準拠し容易なこと。
B　デジタルガンマカメラ【B】
a　本システムは二検出器ガンマカメラを備えた核医学画像診断システムであり、ＳＰＥＣＴ断層撮影により様々な放射性医薬品を利用できること。
b　コリメータは、心臓検査を含んだ各種検査に対応し、交換がスムーズに行えること。
c　患者天台は、昇降が容易な低床式天台であること。
d　操作性に優れ、画像処理及び機能解析が高速かつ容易に行えること。
e　画像データ及び関連情報の転送保管がＤＩＣＯＭ準拠し容易なこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成27年２月６日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課契約係　伊藤　憲興　電話025―227―0970
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成26年12月22日から平成27年２月６日まで。
⑵　交付場所　上記２⑵に同じ。