⑸　調達に必要とされる基本的な要求要件

Ａ　超音波手術装置　超音波を利用し、組織を選択的に超音波振動で破砕・乳化・吸引する機能を有していること。

Ｂ　移動型Ｘ線撮影装置　Ｘ線管は回転陽極であり、モニタは２面でアーム式により、自由に可動できる機能を有していること。

Ｃ　全身麻酔器　麻酔器本体ベンチレータ部、患者回路システム部、気化器部、余剰ガス排出装置、予備用酸素ボンベ搭載部、本体一体型患者呼吸カラーディスプレイ部から構成されていること。

Ｄ　高周波手術装置　モノポーラ・バイポーラの機能を有すること。また、モードには凝固深度や速度を５段階以上の変更が出来る機能を有すること。

Ｅ　脊椎専用移動型手術台　脊椎手術のための椎弓切除・除圧・骨切・前方／後方脊椎固定・脊柱変形矯正において最適な脊椎専用手術テーブルであること。

Ｆ　手術用照明器　光源にＬＥＤを使用しており、主灯160,000lux以上、副灯100,000lux以上の照度を要すること。また、主灯中心にＨＤカメラが搭載できること。

Ｇ　脳外科用移動型手術台　昇降、縦転、横転、背板屈折、縦スライド、横スライド機能を有し、縦スライド・横スライドに関してはフットスイッチにて術者自らが手術台を動かす機能を有していること。

Ｈ　自動洗浄装置　器材、器具、スリッパの洗浄・除染及び乾燥を全自動で行う装置であること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成27年５月15日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒860―8556　熊本市中央区本荘１－１－１　熊本大学医学部附属病院事務部財務・調達ユニット　速瀬あずさ　電話096―373―5938

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成27年３月26日から平成27年５月15日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成27年４月３日14時00分

⑵　開催場所　熊本大学医学部附属病院第二会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Advanced surgery support system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Ultrasonic surgical device　It has a function that utilizes ultrasound to crush, emulsify, aspirate the tissue selectively with ultrasonic vibration.

B　Mobile X-ray imaging device　X-ray tube is a rotating anode, a monitor is arm type and has the ability to freely movable in two planes.

C　General anesthesia unit　That is composed of the anesthesia machine ventilator body part, patient circuit system parts, carburetor parts, excess gas discharge device, spare oxygen cylinder mount, from the body-integrated patient breathing color display unit.

D　High frequency surgical device　It has a function of monopolar-bipolar. Further, its mode has a function of changing out of five or higher rate in solidification depth and speed.

E　Spine-only mobile operating table　That is the best spine surgery-only table in laminectomy, decompress, osteotomy, posterior／anterior spinal fusion, spinal deformity correction for spine surgery.

F　Surgical illuminator　It uses a LED light source, the main light is above 160,000 lux, the illuminance of auxiliary light is 100,000 lux or more. Further, the HD camera can be mounted on the main light area.

G　Mobile operating table for brain surgery　It has a function of lifting, vertical rotation, horizontal rotation, backboard refraction, vertical slide and horizontal slide. And it also has a function that the operator moves the operating table at the foot switch with respect to the vertical slide and horizontal slide.

H　Automatic cleaning equipment　It is a device for performing fully automatic washing, decontamination and drying equipment, instruments, slippers.

⑸　Time limit for the submission of the requested material : 17 : 00　15 May, 2015

⑹　Contact point for the notice : Azusa Hayase, Finance and Procurement Division, Hospital Administration Office, Kumamoto University, 1―1―1 Honjo, Chuo-ku, Kumamoto-shi 860―8556 Japan, TEL 096―373―5938

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 27 年 3 月 26 日

大学共同利用機関法人

情報・システム研究機構長　北川源四郎

◎調達機関番号　416　◎所在地番号　13

○第８号―情報研

１　調達内容

⑴　品目分類番号　71、29

⑵　導入計画物品及び数量　学術情報ネットワーク（ＳＩＮＥＴ５）用　監視運用業務一式

⑶　調達方法　購入等

⑷　導入予定時期　平成27年度１月以降

⑸　調達に必要とされる基本的な要求要件

①　ＳＩＮＥＴ５の監視・運用を行い、受付・入退館・接続作業・障害等に対応すること。

②　監視運用業務の実施場所、ＳＩＮＥＴ５バックボーン接続環境、ＳＩＮＥＴ５監視網接続環境等を用意すること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成27年４月27日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒101―8430 東京都千代田区一ツ橋２－１－２　学術総合センター　大学共同利用機関法人情報・システム研究機構　国立情報学研究所総務部会計課　長谷川美奈　電話03―4212―2053

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成27年３月26日から平成27年４月27日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成27年４月２日14時00分

⑵　開催場所　学術総合センター22階会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the services to be procured : 71, 29

⑵　Nature and quantity of the services to be required : Science Information NETwork (SINET5) 24-hour Monitoring and Operation Service 1 Set

⑶　Type of the procurement : Require

⑷　Basic requirements of the procurement :

①　The procurement should include monitoring and operation of SINET5, and coping with reception of inquiries, entrance／exit management, connection configuring, and failures.