(4) Basic requirements of the procurement :
A The MRI system should have the capability of acquiring images in focused local region with extra high resolutions, and further obtaining those covering total whole body.
B The MRI system should be equipped with large bore magnet and be capable of silent image scanning.
C The MRI system should have optical digital RF transmit and receive facility which may make routine clinical images increase their signal to noise ratio and clarity than ever.
D The MRI system should have the detachable patient table system, easily and safely for patients.
E The MRI system should provide the tools to help obtaining image planning for radiotherapy.
(5) Time limit for the submission of the requested material : 17 : 00　12 January, 2016
(6) Contact point for the notice : Hironao Matsubara, Procurement Section, Accounting Division, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3105

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成27年11月26日
国立大学法人鳥取大学長　豊島　良太
◎調達機関番号　415　◎所在地番号　31
○第２号
1　調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　血管撮影システム一式
(3) 調達方法　借入
(4) 導入予定時期　平成28年度３月以降
(5) 調達に必要とされる基本的な要求要件
A　心臓血管撮影装置
a　循環器領域対応のＣアームバイプレーン装置であること。
b　各アームにＦＰＤ検出器を搭載していること。
c　多画像の表示が可能な大型ＴＶモニターを天吊りで有すること。
d　３次元画像データの取得・再構成が可能であること。
e　最新の循環器用アプリケーションを備えた画像処理ワークステーションを有すること。
f　不整脈治療装置、心腔内電極カテーテル３次元画像装置とのインターフェースを有すること。
g　動画・静止画データの保管及び配信ができること。
h　当院のＨＩＳ、ＲＩＳ及びＰＡＣＳと接続し情報の保存・送受信ができること。
i　保守サービス体制が整っていること。
B　頭部血管撮影装置
a　多目的血管対応のＣアームバイプレーン装置であること。
b　各アームにＦＰＤ検出器を搭載していること。
c　多画像の表示が可能な大型ＴＶモニターを天吊りで有すること。
d　３次元画像データの取得・再構成が可能であること。
e　最新の脳血管用アプリケーションを備えた画像処理ワークステーションを有すること。
f　動画・静止画データの保管及び配信ができること。
g　当院のＨＩＳ、ＲＩＳ及びＰＡＣＳと接続し情報の保存・送受信ができること。
h　保守サービスが整っていること。
C　ＩＶＲ―ＣＴ搭載型血管撮影装置
a　装置は、血管撮影装置とＣＴ装置が含まれた結合型機器であること。
b　血管撮影装置にはＦＰＤ検出器を搭載し、ＣＴ装置には64列以上の検出器を搭載したＩＶＲ支援装置であること。
c　多画像の表示が可能な大型ＴＶモニターを天吊りで有すること。
d　３次元画像データの取得・再構成が可能であること。
e　最新の腹部血管用アプリケーションを備えた画像処理ワークステーションを有すること。
f　動画・静止画データの保管及び配信ができること。
g　当院のＨＩＳ、ＲＩＳ及びＰＡＣＳと接続し情報の保存・送受信ができること。
h　ＩＶＲ治療支援多目的超音波装置を有すること。
i　保守サービスが整っていること。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　平成28年１月15日17時00分（郵送の場合は必着のこと。）
(2) 提供先　〒683―8504 米子市西町36―１ 鳥取大学医学部経理・調達課病院調達係　坂本千亜紀　電話0859―38―7087
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　平成27年11月26日から平成28年１月15日まで。
(2) 交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1) 開催日時　平成27年12月３日13時00分
(2) 開催場所　鳥取大学医学部附属病院会議室(4)
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be rent : Angiography System 1 Set
(3) Type of the procurement : Rent
(4) Basic requirements of the procurement :
A The Cardiac angiography system.
a The system should be a bi-plane angiographic unit equipped with C-arms to support cardiac interventions.
b Each C-arm should be equipped with flat-panel detector system.
c Large-sized TV monitor sets which can display multiple image series should be suspended from a ceiling.
d The system can enable acquisition and reconstruction of three-dimensional data.
e The system should have a workstation in which leading-edge software programs for cardiac studies are installed.
f The system should have interface to equipment for arrhythmia therapy and three-dimensional imaging system intra-cardiac electrode catheter.
g The unit can store and send motion and static images.
h The unit should be connected with HIS, RIS and PACS of the hospital and their information can be transferred each other.
i Maintenance service system for the system is in place.
B The cerebral angiography system.
a The system should be a bi-plane angiographic unit equipped with C-arms to support cerebral intervention especially.
b Each C-arm should be equipped with flat-panel detector system.
c Large-sized TV monitor sets which can display multiple image series should be suspended from a ceiling.
d The system can enable acquisition and reconstruction of three-dimensional data.
e The system should have a workstation in which leading-edge software programs for cerebral angiography studies are installed.
f The unit can store and send motion and static images.
g The unit should be connected with HIS, RIS and PACS of the hospital and their information can be transferred each other.
h Maintenance service system for the system that is in place.