② システムの性能
(A) 多目的デジタルＸ線透視撮影装置であること。
(B) フラットパネルディテクタを搭載し、広範囲の透視、撮影が可能であること。
(C) 寝台が昇降可能であること。
(D) 短時間撮影が出来るＸ線高電圧装置を備えること。
(E) 既設の放射線画像システム及び病院情報システムに接続出来ること。
③ その他
(A) 操作性及び安全性に優れていること。
(B) 将来の機能拡張に対応できること。

2 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提出期限 平成28年２月15日17時00分（郵送の場合は必着のこと。）

(2) 提出先 〒359―8513 埼玉県所沢市並木３－２ 防衛省防衛医科大学校事務局経理部経理課用度第２係 鶴島 織絵 電話04―2995―1877

3 その他 本公表内容は予定であり、変更することがあり得る。

4 Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased : Radiographic and Fluoroscopic X-Ray System 1 Set

(3) Type of the Procurement : purchase

(4) Basic requirements of the Procurement :
① Radiographic and Fluoroscopic X-Ray System should be equipped with every standard and optional attachment.
② Performance of the system :
(A) It must be a multipurpose digital radiographic and fluoroscopic X-ray system.
(B) A flat panel detector must be carried and wide range fluoroscopy and radiography must be possible.
(C) The table must be able to rise and fall.
(D) X-ray high-voltage equipment must perform short-time radiography.
(E) It should be connected with existing RIS, PACS and HIS.
③ Others
(A) It should be safe and easy to be operated.
(B) It should be able to be upgraded in the future.

(5) Time limit for the submission of the requested material : 17 : 00 15, February, 2016

(6) Contact point for the notice : Orie Tsurushima, Procurement Section, Accounting Division, Department of Finance, Administration Bureau, National Defense Medical College, 3―2 Namiki Tokorozawa-shi Saitama 359―8513 Japan TEL 04―2995―1877

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成27年12月21日

防衛省防衛医科大学校事務局
経理部長 田部井貞明

◎調達機関番号 010 ◎所在地番号 11

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量
フラットディテクター型デジタルイメージングシステム 一式

(3) 調達方法 購入等

(4) 導入予定時期 平成28年度第４・四半期以降

(5) 調達に必要とされる基本的な要求要件
① システムの構成 フラットディテクター型デジタルイメージングシステム一式
全標準付属品及び全オプション付属品を備えること。
② システムの性能
(A) 多目的デジタルＸ線透視撮影装置であること。
(B) 高解像度とフルビュー画像表示により、術者が詳細且つ広い視野で正確な画像情報を得る事が出来ること。
(C) 長時間の透視を行ってもオーバーヒートの心配なく装置を使用出来ること。
(D) ＤＳＡ撮影とロードマップを行う事が出来ること。
(E) Ｃアームを電動で動かす事が出来、任意のポジションを記憶してすばやく再生する事が出来ること。
(F) 撮影した画像をハードディスクに100,000枚保存出来、撮影した画像をサーバへ転送出来ること。
③ その他
(A) 操作性及び安全性に優れていること。
(B) 将来の機能拡張に対応できること。

2 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提出期限 平成28年２月15日17時00分（郵送の場合は必着のこと。）

(2) 提出先 〒359―8513 埼玉県所沢市並木３－２ 防衛省防衛医科大学校事務局経理部経理課用度第２係 鶴島 織絵 電話04―2995―1877

3 その他 本公表内容は予定であり、変更することがあり得る。

4 Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased : Mobile C-arm digital imaging system with flat detector 1 Set

(3) Type of the Procurement : purchase

(4) Basic requirements of the Procurement :
① Mobile C-arm digital imaging system with flat detector should be equipped with every standard and optional attachment.
② Performance of the system :
(A) It must be a multipurpose digital radiographic and fluoroscopic X-ray system.
(B) It must get detail, wide and correct images for these specifications.
(C) It must be used for long time against to keep temperature of X-ray tube assembly.
(D) It must prepared digital subtraction angiography and roadmap capability.
(E) C-arm must move by motorized assist and keep memories and replays for several positions.
(F) It must keep memories for 100,000 images to hard disk drive and transfer image dates to image saver.
③ Others
(A) It should be safe and easy to be operated.
(B) It should be able to be upgraded in the future.

(5) Time limit for the submission of the requested material : 17 : 00 15, February, 2016

(6) Contact point for the notice : Orie Tsurushima, Procurement Section, Accounting Division, Department of Finance, Administration Bureau, National Defense Medical College, 3―2 Namiki Tokorozawa-shi Saitama 359―8513 Japan TEL 04―2995―1877

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成27年12月21日

防衛省防衛医科大学校事務局
経理部長 田部井貞明

◎調達機関番号 010 ◎所在地番号 11

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量
手術用顕微鏡システム 一式

(3) 調達方法 購入等

(4) 導入予定時期 平成28年度第４・四半期以降

(5) 調達に必要とされる基本的な要求要件
① システムの構成 手術用顕微鏡システム一式
全標準付属品及び全オプション付属品を備えること。
② システムの性能
(A) 術者用、側視観察助手鏡筒及び対向助手鏡筒を搭載していること。
(B) ドレープ吸引機能を有するフロアスタンドを有すること。