The system must be equipped with a monitoring display whereby the ultrasonograms can be observed outside an examination room.

C　Abdominal-superficial ultrasonograph system must be equipped with the functions of being able to synchronize CT images and to support medical procedures including biopsies. The system must be able to send the utrasonograms to Prime Vita and the existing PACS savers and must be equipped with a monitoring display whereby the ultrasonograms can be observed outside an examination room.

D　Prime Vita system upgrade must be able to be connected to cardiac and abdominal-superficial ultrasonographs and must save still images and cines of the ultrasonograms adequately. It must be connected to the existing PACS savor.

E　Clinical Neurology Network System must be equipped with the latest OS required and must be able to be connected to electroencephalograph and electromyograph. It must also be able to show and save all the past images of electroencephalograms and electromyograms. The images saved must be able to be referred to by the existing Clinical Information System in University of Fukui Hospital.

F　The integrative system must be equipped with the device for quality control of the ultrasonographs, an electric bed for autonomic function tests, and a stool for Master's exercise electrocardiograms.

⑸　Time limit for the submission of the requested material : 17 : 00　8 April, 2016

⑹　Contact point for the notice : Masanori Hasegawa, Procurement Section, Accounting Division, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3147

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年2月22日

国立大学法人島根大学長　服部　泰直

◎調達機関番号　415　◎所在地番号　32

○第４号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ハイブリッドＥＲシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成28年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ）重症患者診療において、患者を移動することなく救急初期診療（初療診療）、ＣＴ検査、緊急手術、血管造影の４つの機能が一つの診療台の上で並行して実施でき治療完結できるハイブリッドＥＲシステムであること。

Ｂ）ＣＴ撮影にあっては、患者台の移動を極力少なくし、ＣＴガントリーが移動できる自走式ガントリーシステムであること。

Ｃ）血管造影はＣアーム型のＩＶＲシステムであること。

Ｄ）ＩＶＲ中のＣＴ撮影が可能であること。

Ｅ）撮影したＣＴ画像やＸ線検査画像および電子カルテを大液晶画面（56インチ高精細モニター以上のもの）に表示して閲覧できること。また、本画面には、ＣＴ　２画像、血管造影画像、エコー、電子カルテ、患者バイタルモニター、血液検査結果などが自由に切り替え表示できること。

Ｆ）患者バイタルモニターのミラーモニターを天吊り型で用意できること。

Ｇ）血管造影とＣＴはいずれも迅速に実施が可能であること。

Ｈ）開胸、開腹、開頭術などの手術を実施するに当たっての手術室環境を用意できること。

Ｉ）ハイブリッドＥＲの広さは十分な広さを確保できること。

Ｊ）手術中のハイブリッドＥＲ室内の様子および手術野のビデオ撮影が可能であること。なお、手術野のビデオ撮影にあっては無影灯と別アームで本システムに干渉しないものを用意できること。また、既存の手術室ビデオ撮影システムと連携が取れること。

Ｋ）ビデオ撮影は簡便な方法（例えば、ボタン一つの操作で撮影が開始できるなど）で撮影の開始、終了が行えること。

Ｌ）血管造影中の映像をビデオ記録として保存できること。

Ｍ）天吊り型で医療機器（エコー、麻酔機、呼吸器など）および電子カルテ端末を設置できること。

Ｎ）照明および無影灯は照度の高いものを用意できること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成28年４月８日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒693―8501 出雲市塩冶町89―１ 国立大学法人島根大学医学部会計課契約第二係　向山　孝行　電話0853―20―2043

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成28年２月22日から平成28年４月８日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成28年２月25日17時00分

⑵　開催場所　島根大学医学部本部棟第一会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Hybrid ER system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A）Hybrid ER system must include an emergency survey and resuscitation, CT scan, operation, and angiographic IVR, which are carried out at the same time without moving the patients.

B）CT-gantry must move in the CT scan without moving the patients.

C）Angiography system must be C-arm type IVR system.

D）CT scan must be done during IVR.

E）Images of CT, X-ray, and electric medical records must be shown on the video panel. This video panel must include 2 CT image, angiography, echo image, workstation image, electric medical records, vital sign monitor, and laboratory data, which must be changed easily.

F）A mirror monitor of patient's vital sign must be hanged from the ceiling.

G）CT scan and IVR must be started immediately.

H）Hybrid ER system must include the operation room for various operation.

I）Hybrid ER room must be enough area.

J）Video of the operation field and in the hybrid ER must be recorded. Video arm must be independent from surgical light devices and must not interferes with IVR system. Video system must cooperate with the existing video system in our operation rooms.

K）Video record must be started by simple method (ex. One push).

L）Image of angiography must be recorded on the video.

M）All medical equipment including echo machine, anesthesia apparatus, and respirator must be hanged from the ceiling.

N）Surgical light devices must be high illuminance.

⑸　Time limit for the submission of the requested material : 17 : 00　8 April, 2016

⑹　Contact point for the notice : Takayuki Mukouyama, Contracting Section 2, Accounting Division, The National University Corporation Shimane University Faculty of Medicine, 89―1 Enya-cho Izumo-shi 693―8501 Japan, TEL 0853―20―2043

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年2月22日

国立大学法人広島大学　分任契約担当職

霞地区運営支援部長　下田　修二