資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年3月16日

国立大学法人新潟大学長　髙橋　　姿

◎調達機関番号　415　◎所在地番号　15

○第４号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　放射線情報管理システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成28年度３月以降

⑸　調達に必要とされる基本的な要求要件

A　放射線情報管理システム（ＲＩＳ）は、病院情報システム（ＨＩＳ）と接続し、患者情報、予約情報、検査オーダ情報、検査実施情報の連携を行うこと。

B　放射線部門業務全体の効率化を実現できること。

C　ユーザーログオン機能を使用して、職員管理によるセキュリティ機能が確保すること。

D　検査機器を管理する機能を備えていること。

E　被曝線量を管理する機能を備えていること。

F　検査機器管理、被曝線量管理システムとの連携を行い、ＲＩＳ端末上で管理システムのＷＥＢ画面を呼び出すことが出来ること。

G　既存の医用画像管理システム（ＰＡＣＳ）、放射線レポートシステム、超音波レポートシステム、動画ネットワークシステムとの連携を行うこと。

H　既存ＲＩＳのデータを移行すること。

I　ＲＩＳは、24時間・365日連続稼動できること。また、電子カルテシステム等のシステムトラブルに対しても、最低限の運用制限で業務が行える仕組みであること。

J　ＲＩＳ導入後も機能の拡張、要望への対応が出来るシステムであること。

K　放射線治療情報管理システム（治療ＲＩＳ）は、病院情報システム（ＨＩＳ）と接続し、患者情報、予約情報、検査オーダ情報、検査実施情報の連携を行うこと。

L　治療計画システムとの接続により、治療計画情報（ＤＩＣＯＭ　ＲＴ　Ｐｌａｎ）、治療画像を取り込むことができること。

M　治療装置との接続により、治療装置への患者情報や受付情報の送信、治療装置からの照射結果情報（ＤＩＯＣＭ　ＲＴ　Ｒｅｃｏｒｄ）の受信が可能であること。

N　他システムからの治療情報の参照を可能とするための治療ＷＥＢ参照機能を備えていること。

O　使用記録簿、方向利用率など治療業務で定型帳票関連及び任意に指定した条件にて蓄積データ集計・ＪＡＳＴＲＯ集計や抽出が可能なこと。

P　ＤＩＣＯＭ　ＲＴ－Ｓｅｒｖｅｒ機能を備えていること。

Q　既存治療ＲＩＳのデータ移行を行うこと。

R　治療ＲＩＳは、24時間・365日連続稼動できること。また、電子カルテシステム等のシステムトラブルに対しても、最低限の運用制限で業務が行える仕組みであること。

S　治療ＲＩＳ導入後も機能の拡張、要望への対応が出来るシステムであること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成28年５月２日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課契約係　野水　忠宏　電話025―227―0970

3　説明書の交付本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成28年３月16日から平成28年５月２日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催本公表に基づく導入説明会を開催する。

⑴　開催日時　平成28年３月24日10時00分

⑵　開催場所　新潟大学医歯学総合病院西診療棟３階　第２会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Radiology Information System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　To exchange patient information, reservation information, order information and so on with hospital information system (HIS).

B　To achieve the improvement of efficiency of the work of Radiology Department.

C　To secure security capability by the staff management using a user logon function.

D　To include a function to manage the inspection equipment.

E　To including a function to manage the radiation dose.

F　To call the WEB screen of the inspection equipment system and radiation dose management system on RIS terminal.

G　To cooperate with existing PACS system, radiology reporting system, ultrasound reporting system, and moving image server.

H　To perform data shift of existing RIS.

I　RIS can operate consecutively for 365 days for 24 hours, and be the structure which can perform duties by a minimum operative limit for the HIS system troubles.

J　To be the expansion of the function. The system can support to a demand after RIS introduction.

K　To exchange patient information, reservation information, order information and so on with hospital information system (HIS).

L　To take in treatment plan information (DICOM RT Plan), a treatment image by the connection with the treatment plan system.

M　The transmission of patient information and the reception desk information to medical treatment equipment, the reception of the irradiation result information (DIOCM RT Record) from medical treatment equipment being possible by connection with the medical treatment equipment.

N　To include a WEB function to enable reference of the treatment information from other systems.

O　Accumulation data count, JASTRO count and extraction being possible on fixed form book vote connection and the condition that they appointed optionally by a use blotter, treatment duties including the direction availability.

P　To include DICOM RT―Server function.

Q　To perform data shift of existing RTIS.

R　RTIS can operate consecutively for 365 days for 24 hours, and be the structure which can perform duties by a minimum operative limit for the HIS system troubles.

S　To be the expansion of the function. The system can support to a demand after treatment RIS introduction.

⑸　Time limit for the submission of the requested material : 17 : 00　2 May, 2016

⑹　Contact point for the notice : Tadahiro Nomizu, Procurement Section, Niigata University Medical and Dental Hospital, 754 Asahimachidori 1-bancho Chuo-ku Niigata-shi 951―8520 Japan, TEL 025―227―0970

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年３月16日

国立大学法人広島大学　分任契約担当職

霞地区運営支援部長　下田　修二

◎調達機関番号　415　◎所在地番号　34

○第５号

1　調達内容

⑴　品目分類番号　26

⑵　導入計画物品及び数量　薬剤業務支援システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成28年度10月以降