(4) Basic requirements of the procurement :
(A) The system should be composed of two Magnetic Resonance Imaging (MRI) systems, and the MRI related indispensable analysis and diagnostic／therapeutic systems.
(B) This system should include the below conditions and requirements.
(a) It should be able to provide good tissue contrast and high time resolution, and also pay attention to patient comfort such as quietness during measurement in addition to the excellent function and quality. To achieve these purposes, it has to have a 1.5 tesla and more magnetic field strength.
(b) It should include the easy operating system providing high functional imaging analysis and diagnostic／therapeutic system used with MRI instruments.
(c) It should equip the system to enhance the functionality of the existing 3 tesla MRI instrument.
(d) It should have the function to connect to the PACS in our hospital and transfer medical images bi-directionally to save and manage all of images.
(5) Time limit for the submission of the requested material : 17 : 00　16 May, 2016
(6) Contact point for the notice : Hideki Ikeda, Supplies, University Hospital, Accounting and Supply Affairs Section, National University Corporation, Tokushima University, 2－50－1 Kuramoto-cho Tokushima-shi 770－8503 Japan, TEL 088－633－9559

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成28年3月28日
国立大学法人徳島大学長　香川　　征
◎調達機関番号　415　◎所在地番号　36
○第４号
1　調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　高度低侵襲医療システム　一式
(3) 調達方法　購入等
(4) 導入予定時期　平成28年度３月以降
(5) 調達に必要とされる基本的な要求要件
A　本システムはHD対応内視鏡システム、手術用超音波診断装置、気腹下施術用システム、赤外光観察内視鏡システム、自動血糖コントロールシステム（人工膵臓）、内視鏡手術用カメラ制御システム及び超音波診断装置を含むこと。
B　ＨＤ対応内視鏡システムは、第２世代ハイビジョンの高画質内視鏡システムであり、高画質化により精度の高い手術手技の提供を行うシステムから構成されること。
C　手術用超音波診断装置はRVS (Real-time Virtual Sonography)、ソナゾイド、エラストグラフィ等の特殊機能を搭載した超音波診断装置であり、術中診断精度の向上を目指すものであること。
D　気腹下施術用システムは気腹圧の増減を最小限に抑えた気腹装置であり、気腹圧を安定させ術者のストレスを低減させるものであること。また、既存の配管設備に問題なく接続できるものであること。
E　赤外光観察内視鏡システムは術中内視鏡手術下にセンチネルリンパ節の同定・血流確認が行える赤外内視鏡システムであり、リンパ節の同定や血流観察により手術精度を向上させるものであること。
F　自動血糖コントロールシステム（人工膵臓）はベッドサイド型持続的血糖監視・微量持続注入システムであり、術中の血糖管理を行い、合併症を低減させるものであること。
G　内視鏡手術用カメラ制御システムは圧縮空気を用いてビデオスコープを固定する保持器システムのことであり、手術件数の増加、術者の負担軽減に寄与するものであること。なお、手術室に既に配管されている圧縮空気のバルブに問題なく接続できるものであること。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　平成28年５月16日17時15分（郵送の場合は必着のこと。）
(2) 提供先　〒770－8503 徳島市蔵本町２－50－１　国立大学法人徳島大学病院経理調達課調達係　池田　英樹　電話088－633－9559
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　平成28年３月28日から平成28年５月16日まで。
(2) 交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1) 開催日時　平成28年４月６日14時30分
(2) 開催場所　国立大学法人徳島大学病院日亜ホールＢｌｕｅ
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be purchased : Advanced Minimally Invasive Medical Care System 1 Set
(3) Type of the procurement : Purchase
(4) Basic requirements of the procurement :
A　This system includes HD compliant laparoscopic system, intraoperative ultrasound diagnosis system, gas feeding system for laparoscopic surgery, infrared light observation system for laparoscopic surgery, auto glycemic control system (artificial pancreas), laparoscope control system for laparoscopic surgery and percutaneous ultrasound diagnosis system.
B　HD compliant laparoscopic system is defined as the 2nd generation HD high resolution laparoscope system, which achieves the surgical technique of high precision derived by the high resolution laparoscopic camera system.
C　Intraoperative ultrasound diagnosis system is defined as the diagnostic ultrasound imaging machine with special equipment such as real-time virtual sonography (RVS), Sonazoid US and elastography, which achieves the improvement of the diagnostic precision intraoperatively.
D　Gas feeding system for laparoscopic surgery is defined as the gas feeding device which can achieve minimal change of the gas pressure , can achieve stable gas pressure and can decrease operators' stress. Moreover, this system can be attached to the existing pipes plumbing.
E　Infrared light observation system for laparoscopic surgery is defined as the infrared light camera system which can detect sentinel lymphonodes and can check blood flow under intraoperative laparoscope, which leads the improvement of the operative precision.
F　Auto-glycemic control system (artificial pancreas) is defined as the continuous blood glucose monitoring and continuous microinjection system which controls the intraoperative blood glucose level, and minimizes operative complications.
G　Laparoscope control system for laparoscopic surgery is defined as the flexible-type camera holder moved by compressed air which increases number of surgical cases and can decrease operators' stress. Moreover, this system can be attached to the existing pipes plumbing.
(5) Time limit for the submission of the requested material : 17 : 15　16 May, 2016
(6) Contact point for the notice : Hideki Ikeda, Supplies, University Hospital, Accounting and Supply Affairs Section, National University Corporation, Tokushima University, 2—50—1 Kuramoto-cho Tokushima-shi 770—8503 Japan, TEL 088—633—9559

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成28年3月28日
国立大学法人熊本大学
契約責任者　理事　西川　　泉