H　It is to be considered that an effective automation of operation management of the system economizes on manpower.

I　It is to be considered that the system has the distinction of energy conservation, space-saving, temperature control and low noise level.

J　The maintenance and technical services should be promptly and precisely provided for the system, in case of a failure of the system.

⑸　Time limit for the submission of the requested material : 17 : 00　2 May, 2016

⑹　Contact point for the notice : Hiroyuki Okaniwa, General Affairs Section, Administration Division, Faculty of Dentistry and Dental Hospital, Tokyo Medical and Dental University, 1―5―45 Yushima Bunkyo-ku Tokyo 113―8549 Japan, TEL 03―5803―5410

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年4月1日

国立大学法人浜松医科大学

理事　前田　　広

◎調達機関番号　415　◎所在地番号　22

○第１号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　核医学画像診断システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成28年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　Ｘ線ＣＴを搭載したＳＰＥＣＴ－ＣＴ装置であること。

Ｂ　最適な診断を可能とするコリメータを有していること。

Ｃ　脳、心臓領域、及び全身検査に対応可能であること。

Ｄ　画像解析用ワークステーションを有していること。

Ｅ　核医学画像診断支援機能を有していること。

Ｆ　既存の放射線情報システム（ＲＩＳ）、医用画像管理システム（ＰＡＣＳ）、及び３Ｄ画像処理システムと連携接続可能であること。

Ｇ　システムを設置するための工事を含むこと。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成28年５月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒431―3192　浜松市東区半田山１―20―１　浜松医科大学会計課専門職員（契約担当）　大保　智則　電話053―435―2148

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成28年４月１日から平成28年５月17日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成28年４月11日13時30分

⑵　開催場所　浜松医科大学管理棟２階第二会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Nuclear Medicine Image Diagnosis System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The required system should be a SPECT gamma camera equipped with X-ray CT.

B　The required system should have several collimators that enable us to get optimal images.

C　The required system should be applicable for brain, cardiac and whole body studies.

D　The required system should have a workstation for image analysis.

E　The required system should have a function to assist a diagnostic procedure in nuclear medicine.

F　The required system should have an interface to communicate with RIS, PACS, and 3D image processing device installed in the department.

G　The cost of construction and installation of the system should be included.

⑸　Time limit for the submission of the requested material : 17 : 00　17 May, 2016

⑹　Contact point for the notice : Tomonori Obo, Expert Official, Accounting Section, Hamamatsu University School of Medicine, 1―20―1 Handayama Higashi-ku Hamamatsu-shi 431―3192 Japan, TEL 053―435―2148

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年4月1日

国立大学法人鹿児島大学

契約担当役理事　永井　義美

◎調達機関番号　415　◎所在地番号　46

○第１号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　Ｘ線ＣＴ装置　一式

⑶　調達方法　借入

⑷　導入予定時期　平成28年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　Ｘ線ＣＴ装置は全身用であること。

Ｂ　逐次近似画像再構成法を有すること。

Ｃ　造影剤自動注入装置を有すること。

Ｄ　本システムによる診療を安全かつ円滑に行う上で、必要不可欠な附属周辺機器を有すること。

Ｅ　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、読影システム、及び画像ネットワーク（ＰＡＣＳ）と接続可能であること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成28年５月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒890―8520 鹿児島市桜ケ丘８―35―１　鹿児島大学医学部・歯学部附属病院管理課調達係　吉井あけみ　電話099―275―5045

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成28年４月１日から平成28年５月17日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成28年４月15日16時00分

⑵　開催場所　鹿児島大学医学部・歯学部附属病院管理棟第三会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be rent : High-performance Computed tomography (CT) system 1 Set

⑶　Type of the procurement : Rent

⑷　Basic requirements of the procurement :

A　The CT system must be for whole-body imaging.

B　The CT system must be equipped with successive approximations image reconstruction method.

C　The CT system must be equipped with an automated contrast injector.

D　The CT system must have peripheral devices that facilitate smooth and safe operation.

E　The CT system must be able to connect with the existent HIS, RIS, image interpreting system and PACS.