2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成28年7月8日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒431―3192　浜松市東区半田山1―20―1　浜松医科大学会計課専門職員（契約担当）　大保　智則　電話053―435―2148
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成28年5月23日から平成28年7月8日まで。
⑵　交付場所　上記2⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成28年5月31日14時00分
⑵　開催場所　浜松医科大学管理棟2階第二会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : A suite of physiological function test equipments system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must be equipped with three multi-function electrocardiograph.
B　The system must be epuipped with a long-hours monitoring system for electrocardiograph, which consists of electrocardiographs and its analysis devices.
C　The system must have the exercise load system. The exercise load is a type of treadmill.
D　The system must be equipped with an ultrasonic diagnostic apparatus (cardiac ultrasound dedicated machine).
①　The ultrasonic diagnostic apparatus shall be available for the highly precise imaging.
②　The current function and analysis programs are available and excellent function probes (a sector probe, a linear probe, a pediatric coronary probe, a 3D probe, a transesophageal 3D probe and a 3D analysis software) should be prepared.
E　The system must be equipped with an ultrasonic diagnostic apparatus (abdominal ultrasound dedicated machine).
①　The ultrasonic diagnostic apparatus shall be provided with high specification probes like a convex probe, a micro convex, a linear probe, a sector probe and a 3D probe, for each purpose.
②　Highly precise ultrasonic diagnosis for abdominal organs can be performed, and biopsy with puncture adapters is available.
③　At examination for kidney or prostate, precise biopsy can be done with ultrasonic guidance.
F　The system must be equipped with an ultrasonic diagnostic apparatus (general-purpose machine).
①　The ultrasonic diagnostic apparatus shall be provided with high specification probes like convex, linear, micro linear, sector and 3D probes.
②　At examination for thyroid and mammary glands, precise biopsy can be done with ultrasonic guidance.
G　The system must have a respiratory function testing apparatus.This device must be able to measure vital capacity and forced vital capacity
H　The device should have an interface to communicate with physiology inspection system (named Prime Vita) installed in the department.
I　The cost of construction and installation of the system should be included.
J　All devices must be able to be used at a voltage of 101 ± 6V.
K　Next system and cooperation connection must be possible
⑸　Time limit for the submission of the requested material : 17 : 00　8 July, 2016
⑹　Contact point for the notice : Tomonori Obo, Expert Official, Accounting Section, Hamamatsu University School of Medicine, 1―20―1 Handayama Higashi-ku Hamamatsu-shi 431―3192 Japan, TEL 053―435―2148

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成28年5月23日
国立大学法人香川大学長　長尾　省吾
◎調達機関番号　415　◎所在地番号　37
○第1号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　内視鏡システム一式
⑶　調達方法　借入
⑷　導入予定時期　平成28年度2月以降
⑸　調達に必要とされる基本的な要求要件
A　4台以上の特殊光観察可能なハイビジョン消化器内視鏡システムであること。
B　動画及び静止画をフルハイビジョンで記録可能な記録装置を有すること。
C　消化器内視鏡検査に対応する送水ポンプを有すること。
D　上部消化管内視鏡検査に対応可能なビデオスコープを有すること。
E　下部消化管内視鏡検査に対応可能なビデオスコープを有すること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成28年7月8日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒761―0793　香川県木田郡三木町大字池戸1750―1　香川大学医学部管理課用度第二係　寒川　貴司　電話087―891―2028
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成28年5月23日から平成28年7月8日まで。
⑵　交付場所　上記2⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成28年5月30日11時00分
⑵　開催場所　香川大学医学部管理棟小会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Endoscope system 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　There must be more than 4 sets of high vision imaging system which can use special light.
B　There must be a memory video system which can take high vision movie and stop motions.
C　There must be a water pump which can be used at gastrointestinal endoscopy.
D　There must be a video scope which can be used at gastrointestinal endoscopy.
E　There must be a video scope which can be used at colonoscopy.
⑸　Time limit for the submission of the requested material : 17 : 00　8 July, 2016
⑹　Contact point for the notice : Kangawa Takasi, Procurement Section, Accounting Division, Kagawa University, 1750―1 Oaza Ikenobe Miki-cho Kita-gun Kagawa 761―0793 Japan, TEL 087―891―2028

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成28年5月23日
国立研究開発法人情報通信研究機構
契約担当理事　黒瀬　泰平