E　The system should be capable of isotropic volume CT scan of the whole body.
F　The system should be capable of fast data transmission of a large amount of digital images.
G　The system should work in cooperation with hospital information system (HIS) and radiology information system (RIS).
H　The system should work in a filmless environment including picture archiving and communication system (PACS).
I　Must have a workstation or workstations for image post-processing and analysis.
J　Must have image viewing and reporting system.
⑸　Time limit for the submission of the requested material : 17 : 00　22 July, 2016
⑹　Contact point for the notice : Minoru Hagiwara, Procurement Section, Accounting Division, the University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成28年6月6日
国立大学法人東京大学総長　五神　　真
◎調達機関番号　415　◎所在地番号　13

○第４号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　高磁場ＭＲシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成28年度３月以降
⑸　調達に必要とされる基本的な要求要件
A　全身のＭＲ画像が撮像可能であること。
B　各種のＲＦコイルを有すること。
C　画像再構成及び画像処理が高速で行えること。
D　ＤＩＣＯＭ対応のインターフェイスを有すること。
E　病院情報システム及び放射線情報システムと連動すること。
F　放射線画像情報システムと連動すること。
G　画像解析用ワークステーションおよび画像解析システムを有すること。
H　読影用端末システムを有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成28年７月22日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒113―8655 東京都文京区本郷７－３－１　東京大学医学部附属病院事務部管理課経理チーム　萩原　　稔　電話03―5800―8616
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成28年６月６日から平成28年７月22日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成28年６月13日10時00分
⑵　開催場所　東京大学医学部附属病院管理課会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : High field MR system 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Must be capable of whole body magnetic resonance imaging.
B　Must have various RF coils for whole body imaging.
C　Must have high-speed image processing ability and high-speed image reconstruction ability.
D　Must have an interface for DICOM.
E　The system should work in cooperation with hospital information system (HIS) and radiology information system (RIS).
F　The system should work in a filmless environment including picture archiving and communication system (PACS).
G　Must have a workstation or workstations for image post-processing and analysis.
H　Must have image viewing and reporting system.
⑸　Time limit for the submission of the requested material : 17 : 00　22 July, 2016
⑹　Contact point for the notice : Minoru Hagiwara, Procurement Section, Accounting Division, the University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成28年６月６日
国立大学法人東京大学総長　五神　　真
◎調達機関番号　415　◎所在地番号　13

○第５号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　放射線治療CTシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成28年度３月以降
⑸　調達に必要とされる基本的な要求要件
A　空間分解能に優れた全身のデジタルＸ線画像が得られること。
B　濃度分解能に優れた全身のデジタルＸ線画像が得られること。
C　時間分解能に優れた全身のデジタルＸ線画像が得られること。
D　画像の再構成が高速に行えること。
E　高精度な位置決め用シミュレーター機能を有すること。
F　大量のデジタル画像情報を高速に転送できること。
G　病院情報システム及び放射線情報システムと連動すること。
H　放射線画像情報システムと連動すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成28年７月22日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒113―8655 東京都文京区本郷７－３－１　東京大学医学部附属病院事務部管理課経理チーム　萩原　　稔　電話03―5800―8616
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成28年６月６日から平成28年７月22日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成28年６月13日10時30分
⑵　開催場所　東京大学医学部附属病院管理課会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Radiation therapy CT system 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system should provide digital cross-sectional images of the whole body with excellent spatial resolution.
B　The system should provide digital cross-sectional images of the whole body with excellent density resolution.
C　The system should provide digital cross-sectional images of the whole body with excellent temporal resolution.
D　The system should be capable of rapid reformation of scanned images.