4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成29年３月16日14時00分
⑵　開催場所　東北大学病院医学部３号館７階共用会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Ultrasonographic Diagnostic System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A）This system consists of three ultrasonographic diagnostic devices, one being for use in the Advanced Critical Care and Emergency Center, one being for use with laparotomy in the Surgery Dept., and one being for endoscopy in the Surgery Dept.
B）The ultrasonographic diagnostic device for use in the Advanced Critical Care and Emergency Center has a B-mode, an M-mode, and a Doppler mode, and support contrast-enhanced ultrasound imaging using perflubutane microbubbles, etc. It can be used for rapid diagnosis of the patients at the time of arriving at emergency center, supports various probes and frequencies suitable for various observational purposes, and can be used for observation and analysis of the properties, etc., of organs and tissues.
C）The ultrasonographic diagnostic device for use with laparotomy in the Surgery Dept. has a B-mode, an M-mode, and a Doppler mode, and support contrast-enhanced ultrasound imaging using perflubutane microbubbles, etc. It has functions that enable the following, primarily for intraoperative echographic diagnosis :
(i)　Detailed assessment of preoperative examination results, such as confirmation of relationships with surrounding organs, and accurate confirmation of lesion positions.
(ii)　Detection of lesions that were unknown before surgery, including verification of deep lesions, and confirmation of the properties of lesions discovered, by palpation, etc., during surgery.
(iii)　Evaluation of the blood flow in organs, involving ascertainment of the blood flow volume, speed, and waveform. The device is equipped with probes and frequencies that are suitable for various observational purposes, and can be used for observation and analysis of the properties, etc., of organs and tissues.
D）The ultrasonographic diagnostic device for use for endoscopy in the Surgery Dept. can be connected to an endoscopy probe that can be inserted from a 12-mm trocar used in laparoscopic surgery, has a B-mode and a Doppler mode, and support contrast-enhanced ultrasound imaging using perflubutane microbubbles, etc. It has functions that enable the following, primarily for intraoperative echographic diagnosis :
(i)　Detailed assessment of preoperative test results, such as confirmation of relationships with surrounding organs, and accurate confirmation of lesion positions.
(ii)　Detection of lesions that were unknown before surgery, including verification of deep lesions, and confirmation of the properties of lesions discovered, by palpation, etc., during surgery.
(iii)　Evaluation of the blood flow in organs, involving ascertainment of the blood flow volume, speed, and waveform. The device is equipped with deep probes and frequencies that are suitable for various observational purposes, and can be used for observation and analysis of the properties, etc., of organs and tissues.
E）The three ultrasonographic diagnostic devices, for use in the Advanced Critical Care and Emergency Center, for use with laparotomy in the Surgery Dept., and for endoscopy in the Surgery Dept., can all be connected to, send information to, and be stored with, existing hospital information systems.
⑸　Time limit for the submission of the requested material : 17 : 00　21 April, 2017
⑹　Contact point for the notice : Hirotaka Kikuchi, Procurement Services Office, Finance Department, Tohoku University, 2―1―1 Katahira Aoba-ku Sendai-shi 980―8577 Japan, TEL 022―217―4869

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成29年３月６日
国立大学法人岡山大学学長　森田　　潔
◎調達機関番号　415　◎所在地番号　33
○第２号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　周術期部門管理システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成29年度９月以降
⑸　調達に必要とされる基本的な要求要件
A　生体情報モニタリングシステム
a　本システムはセントラルモニタ、ベッドサイドモニタ及び送信機により構成されること。
b　重症系患者の生体情報を集中的に監視できること。
c　既存の病院医療情報システムと連携できること。
d　電子カルテに接続し、データの保存ができること。
e　外部機器からのデータを取り込みできること。
B　重症系患者管理システム
a　電子カルテシステムとネットワーク構築ができること。
b　ネットワークサーバーおよび診療端末などの関連ハードウエアおよび付属装置を備えていること。
c　生体情報モニタリングシステムと連携が可能であること。
d　長期的に、品質保証、品質管理に対応できること。
e　得られた各種データを記録し必要に応じて検索することができ、患者の治療、医学教育、病院経営改善に有効に利用できること。
f　進化する各種生体情報モニタからのデータ取り込みに速やかに対応できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成29年４月21日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒700―8558　岡山市北区鹿田町２－５－１　国立大学法人岡山大学　岡山大学病院経営・管理課調達グループ政府調達担当　上廻　真司　電話086―235―7540
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成29年３月６日から平成29年４月21日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成29年３月13日14時00分
⑵　開催場所　国立大学法人岡山大学医学部管理棟小会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Perioperative care management system 1 Set