⑷ Basic requirements of the procurement :
A The central monitor must be to meet the following requirements.
a The central monitor should be deployed in each staff station, and information on the bedside monitor can be monitored intensively.
b System cooperation with hospital information system operating at our hospital is possible.
B The bedside monitor must be to meet the following requirements.
a The bedside monitor should be able to measure waveforms and display numerical values by measuring electrocardiogram, respiration, SpO2, noninvasive blood pressure, blood flow pressure, body temperature etc.
b The bedside monitor sets parameters and alarms for children and newborn patients having a function to determine.
c Have sensors for children and newborns.
C The pulse oximeter must be to meet the following requirements.
a The pulse oximeter must be a single unit.
b The pulse oximeter should have the function to measure SpO2.
c Connect with the bedside monitor and work with the function.
⑸ Time limit for the submission of the requested material : 17 : 00　26 January, 2018
⑹ Contact point for the notice : Risako Abe, Procurement Section, Accounting Division, the University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 29 年 12 月 11 日
国立大学法人東京大学総長　五神　　真

◎調達機関番号　415　◎所在地番号　13
○第７号
1　調達内容
⑴ 品目分類番号　22、31
⑵ 導入計画物品及び数量　生体情報モニタリングシステム　一式
⑶ 調達方法　購入等
⑷ 導入予定時期　平成30年度12月以降
⑸ 調達に必要とされる基本的な要求要件
A セントラルモニタは以下の要件を満たすこと。
a セントラルモニタは各スタッフステーションに配備し、ベッドサイドモニタの情報を集中的に監視可能なこと。
b 本院で稼働している病院情報システムとのシステム連携が可能であること。
B ベッドサイドモニタは以下の要件を満たすこと。
a ベッドサイドモニタは、心電図､呼吸、SpO２、非観血血圧、観血血圧、体温等を測定し波形表示及び数値表示ができること。
b ベッドサイドモニタは､患者タイプ(成人他）に合わせて、パラメータやアラーム設定を決定できる機能を有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴ 資料等の提供期限　平成30年１月26日17時00分（郵送の場合は必着のこと。）
⑵ 提出先　〒113―8655 東京都文京区本郷７－３－１　東京大学医学部附属病院事務部管理課経理チーム　大内　啓彰　電話03―5800―8616
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴ 交付期間　平成29年12月11日から平成30年１月26日まで。
⑵ 交付場所　上記２⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴ Classification of the products to be procured : 22, 31
⑵ Nature and quantity of the products to be purchased : The Patient Monitor 1 set
⑶ Type of the procurement : Purchase
⑷ Basic requirements of the procurement :
A The central monitor must be to meet the following requirements.
a The central monitor should be deployed in each staff station, and information on the bedside monitor can be monitored intensively.
b System cooperation with hospital information system operating at our hospital is possible.
B The bedside monitor must be to meet the following requirements.
a The bedside monitor should be able to measure waveforms and display numerical values by measuring electrocardiogram, respiration, SpO2, noninvasive blood pressure, blood flow pressure, body temperature etc.
b The bedside monitor sets parameters and alarms for adult patients having a function to determine.
⑸ Time limit for the submission of the requested material : 17 : 00　26 January, 2018
⑹ Contact point for the notice : Hiroaki Ouchi, Procurement Section, Accounting Division, the University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 29 年 12 月 11 日
国立大学法人東京大学総長　五神　　真

◎調達機関番号　415　◎所在地番号　13
○第８号
1　調達内容
⑴ 品目分類番号　22、31
⑵ 導入計画物品及び数量　保育器　一式
⑶ 調達方法　購入等
⑷ 導入予定時期　平成30年度12月以降
⑸ 調達に必要とされる基本的な要求要件
A 開放型保育器は以下の要件を満たすこと。
a 遠赤外線ヒーターの輻射熱等による児の保温ができること。
b 児の蘇生や外科的処置が可能な機能を有すること。
c 薬剤で清拭できること。
B 閉鎖型保育器は以下の要件を満たすこと。
a 庫内の温度・湿度・酸素濃度の設定ができ、庫内の環境をモニタリングできること。
b 庫内に収容された児が外からでも観察できること。
c 庫内に収容された児の処置が容易にできること。
d 児の睡眠を妨げないよう静音性能を有すること。
e 児の体重計測ができること。
f 薬剤で清拭できること。
C 搬送用保育器は以下の要件を満たすこと。
a 庫内の温度・湿度・酸素濃度の設定ができ、庫内の環境をモニタリングできること。
b 庫内に収容された児が外からでも観察できること。
c 搬送できるよう専用ストレッチャーを有すること。
d 保育器本体とストレッチャーが脱着できること。
e バッテリーを有し駆動できること。
f 薬剤で清拭できること。
D 光線治療器は以下の要件を満たすこと。
a 新生児高ビリルビン血症に対する光線療法に使用できること。
b 保育器に収容された児に治療用の光線を照射できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴ 資料等の提供期限　平成30年１月26日17時00分（郵送の場合は必着のこと。）
⑵ 提出先　〒113―8655 東京都文京区本郷７－３－１　東京大学医学部附属病院事務部管理課経理チーム　石川　亜実　電話03―5800―8616