3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成29年12月11日から平成30年1月26日まで。
⑵　交付場所　上記2⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Infant Incubator 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The open type infant incubator must be to meet the following requirements.
a　The unit must be able to warm an infant by means of radiation heat of far infrared heater, etc.
b　The unit must have functions capable of resuscitation measures and surgical treatments for an infant.
c　The unit must be able to be cleaned by medical fluid.
B　The closed-type infant incubator must be to meet the following requirements.
a　The temperature, humidity and oxygen concentration in the unit must be able to be set and the environment in the unit must be able to be monitored.
b　The condition of an infant in the unit must be able to be observed from outside.
c　An infant in the unit must be able to be easily treated.
d　The unit must have silent performance so as not to disturb a sleep of an infant.
e　The unit must have a function to measure a body weight of an infant.
f　The unit must be able to be cleaned by medical fluid.
C　The transport incubator must be to meet the following requirements.
a　The temperature, humidity and oxygen concentration in the unit must be able to be set and the environment in the unit must be able to be monitored.
b　The condition of an infant in the unit must be able to be observed from outside.
c　The unit must have a dedicated stretcher for transportation.
d　The main unit of the incubator must be able to be detached from the stretcher.
e　The unit must have a battery and be powered by the battery.
f　The unit must be able to be cleaned by medical fluid.
D　The phototherapy unit must be to meet the following requirements.
a　The unit must be able to be used for phototherapy for the treatment for neonatal hyperbilirubinemia.
b　The unit must be able to irradiate a therapeutic light to an infant in the incubator.
⑸　Time limit for the submission of the requested material : 17 : 00　26 January, 2018
⑹　Contact point for the notice : Ami Ishikawa, Procurement Section, Accounting Division, the University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成29年12月11日
国立大学法人東京大学総長　五神　真
◎調達機関番号　415　◎所在地番号　13
○第９号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　分娩監視システム一式
⑶　調達方法　購入等
⑷　導入予定時期　平成30年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　分娩監視装置は以下の要件を満たすこと。
a　外側法による胎児心拍数と子宮収縮を測定することができ、本体に数値が表示されること。また記録紙上に印刷することができること。
b　設定した警報値に対して、アラートが出ること。
c　セントラルモニタへデータが反映できること。
d　超音波トランスデューサと陣痛トランスデューサを有すること。
e　機種によっては双胎および品胎の胎児心拍数を測定することができ、多胎の胎児心拍数を記録紙上に印刷することができること。
f　機種によっては内側法による胎児心拍数および胎児心電図、子宮内圧を測定することができること。
g　機種によっては血圧、血中酸素飽和度が測定でき、記録紙上に印刷することができること。
B　分娩監視セントラルモニタは以下の要件を満たすこと。
a　分娩監視セントラルモニタは、スタッフステーションに配備し、分娩監視装置の情報を集中的に監視可能なこと。
b　本院で稼働している病院情報システムとのシステム連携が可能であること。
c　胎児心拍数、母体心拍・脈拍数、母体非観血血圧数、母体動脈血酸素飽和度の各測定値が、上限値を超えた場合もしくは下限値を下回った場合に報知すること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年1月26日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒113―8655 東京都文京区本郷7－3－1　東京大学医学部附属病院事務部管理課経理チーム　中西　雅通　電話03―5800―8616
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成29年12月11日から平成30年1月26日まで。
⑵　交付場所　上記2⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : The fetal monitor System 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The fetal monitor must be to meet the following requirements.
a　Measure the fetal heart rate and uterine contractions by the external measuring methods and those numeric values measured shall be displayed on the monitor. Moreover, it can print on recording paper.
b　Give off alarms at preset alarm values.
c　Reflect the data in central station monitors.
d　Come with an ultrasound transducer and a Toco transducer.
e　Depending on the model, measure the fetal heart rates of twins and triplets, and the fetal heart rates of such multiplets shall be printable on the sheet of recording paper.
f　Depending on the model, measure the fetal heart rate, fetal electrocardiogram and intrauterine pressure by the internal measuring methods.
g　Depending on the model, measure the blood pressure and the blood oxygen saturation, print it out on the recording paper.
B　The central station monitor for bidding must be to meet the following requirements.
a　The central station monitor for bidding is deployed to the staff station, Ability to intensively monitor information on the fetal monitor.