⑸　福井県、滋賀県、京都府、大阪府、兵庫県、奈良県及び和歌山県　近畿地方整備局総務部契約課（〒540―8586 大阪府大阪市中央区大手前１－５－44大阪合同庁舎第一号館（電話06―6942―1141））

⑹　鳥取県、島根県、岡山県、広島県及び山口県　中国地方整備局総務部契約課（〒730―8530 広島県広島市中区上八丁堀６－30広島合同庁舎第二号館（電話082―221―9231））

⑺　徳島県、香川県、愛媛県及び高知県　四国地方整備局総務部契約課（〒760―8554 香川県高松市サンポート３－33高松サンポート合同庁舎（電話087―851―8061））

⑻　福岡県、佐賀県、長崎県、熊本県、大分県、宮崎県、鹿児島県及び沖縄県　九州地方整備局総務部契約課（〒812―0013 福岡県福岡市博多区博多駅東２－10－７福岡第二合同庁舎（電話092―471―6331））

招　請

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成30年２月16日

独立行政法人国立病院機構

大阪医療センター院長　是恒　之宏

◎調達機関番号　597　◎所在地番号　27

○第 08803 号（№08803）

１　調達内容

⑴　品目分類番号　22、32

⑵　導入計画物品及び数量　臨床検査部門機器賃貸借複合契約　一式

⑶　調達方法　借入等

⑷　導入予定時期　平成30年12月まで。

⑸　調達に必要とされる基本的な要求要件　臨床検査機器及びその関連試薬等を総合的に管理・運用することにより、患者の治療に寄与することを目的とする。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成30年４月３日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒540―0006 大阪府大阪市中央区法円坂２－１－14　独立行政法人国立病院機構大阪医療センター契約係長　立堀　善久　電話06―6942―1331　内線8035

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成30年２月16日から平成30年４月３日まで。

⑵　交付場所　上記２⑵に同じ。

４　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

５　Summary

⑴　Classification of the products to be leased : 22, 32

⑵　Nature and quantity of the products to be leased : Inspection machinery compound leased 1 set

⑶　Type of the procurement : Lease

⑷　Time-limit for the submission of the requested material : 5 : 00 P.M April 3, 2018

⑸　Contact point for the notice : Yoshihisa Tachibori, Chief clerk, Accounting Division, National Hospital Oraganization Osaka National Hospital, 2―1―14 Hoenzaka Chuo-ku Osaka-shi, Osaka-fu 540―0006, Japan. TEL 06―6942―1331 ext. 8035

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成30年２月16日

国立大学法人北海道大学総長　名和　豊春

◎調達機関番号　415　◎所在地番号　01

○第４号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ＳＰＥＣＴ－ＣＴシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成31年度６月以降

⑸　調達に必要とされる基本的な要求要件

①　全身の機能診断のためのＳＰＥＣＴ画像およびＣＴ画像を取得できる機能を有すること。

②　核種エネルギーに合わせた専用のコリメータを有し、空間分解能・感度に優れていること。

③　骨シンチグラフィ、脳血流シンチグラフィ、心筋血流シンチグラフィ、ヨード131シンチグラフィの画像を取得できる機能を有すること。

④　ＣＴ画像による減弱・散乱補正を行うことにより集積率などの定量値を測定できること。

⑤　ＣＴ画像は診断ができる十分な空間分解能を持つ画質を有すること。

⑥　診断のための各種画像解析および機能解析ができる機能を有すること。

⑦　画像情報の保管及び配信ができること。

⑧　当院の放射線情報システム・病院情報システムと接続し、画像情報・オーダ情報の通信ができること。

⑨　放射線読影システムは、医用放射線画像の表示および読影ができる機能を有すること。

⑩　放射線測定器は、電離箱式・ＧＭ・シンチレーションサーベイメータより構成され、空間線量測定および表面汚染測定ができること。

⑪　医療機器として安全性が考慮されていること。

⑫　本件調達物品が稼働するために必要な設置場所の改修を行うこととし、遮へい計算により必要なＸ線防護を施すこと。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成30年４月５日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒060―0813 札幌市北区北13条西６丁目　北海道大学病院管理課用度第一係　岩部　　順　電話011―706―7401

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成30年２月16日から平成30年４月５日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成30年２月23日10時00分

⑵　開催場所　北海道大学病院管理課入札室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : SPECT-CT system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

①　The SPECT-CT system must be acquired with the SPECT images and CT images of a whole body for functional diagnostics.

②　The SPECT-CT system must be equipped with a dedicated collimator from low to high energy, and have high spatial resolution and sensitivity.

③　The SPECT-CT system must allow acquisition of bone scintigraphy, brain perfusion scintigraphy, myocardial perfusion scintigraphy, and iodine-131 scintigraphy.

④　The SPECT-CT system must enable quantitative measurements with attenuation and scatter corrections using CT images.

⑤　The SPECT-CT system must have sufficient spatial resolution of CT images for diagnostic use.

⑥　The SPECT-CT system must enable image analysis, including functional analysis for diagnosis.

⑦　The SPECT-CT system must be capable of archiving and distributing the image data.

⑧　The SPECT-CT system must have connections with the hospital's radiology and hospital information systems, and must allow communication of images and order information.

⑨　The medical image reading system must enable display of medical radiographic images for diagnosis.

⑩　The radiation measuring instruments that consist of ionization chamber, Geiger-Mueller survey meter, and scintillation survey meter must be capable of measurements of a spatial radiation dose and radioactive surface contamination.

⑪　The SPECT-CT system must be safe enough for application as medical equipment.