⑨　The bidder should provide peripheral equipment that is effective for managing machines and devices
⑩　The bidder should have an organization which provides quick user support and adequate maintenance of the system
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M 6, July, 2018
⑹　Contact point for the notice : Masataka Tsujimoto, Finance and Accounting Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext. 2982

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年5月22日
国立研究開発法人
国立循環器病研究センター
理事長　小川　久雄

◎調達機関番号　820　◎所在地番号　27
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　血管撮影装置　1式（新センター　カテーテル室⑨に該当）
⑶　調達方法　購入等
⑷　導入予定時期　平成31年1月末日
⑸　調達に必要とされる基本的な要求要件
①　正面・側面の2方向から透視、撮影が可能なバイプレーンシステムであること。
②　慢性肺動脈血栓性肺高血圧症の確定診断、治療が安全に行える装置であること。
③　患者を移動することなく、頭部から下肢まで全身を、検査可能な機能を有すること。
④　X線変換効率、濃度分解能が高い平面検出器（フラットパネルディテクタ）搭載装置であること。
⑤　フラットパネルディテクタは、右心カテーテルによる診断、治療が行えること。
⑥　デジタル画像処理システムは、非血管系治療を支援する機能や血管解析機能を有すること。
⑦　多画像の表示が可能な大型液晶ディスプレイを有すること。
⑧　3次元画像データの取得、再構成が可能であること。
⑨　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、医用画像管理システム（ＰＡＣＳ）、読影システムと接続可能であること。
⑩　装置および機器の運営管理に有用な周辺機器を備えていること。
⑪　装置の維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料の提供期限　平成30年7月6日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台5―7―1　国立研究開発法人国立循環器病研究センター　財務経理部契約第一係　辻本　真孝　電話06―6833―5012（内線）2982
メールアドレス　Sounding@ml.ncvc.go.jp
3　導入説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年5月22日から平成30年7月6日まで。
⑵　交付方法　2⑵に示す場所にて直接交付するか、若しくは電子メールにて交付する。電子メールでの交付を受ける場合は2⑵に示すメールアドレスに調達件名・会社名・担当者名・連絡先を入力し、参加希望の旨メール送信すること。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更する事があり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : X-ray Angiography system 1 set (should be installed catheter room ⑨ in the new center hospital)
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The system should a bi-plane system which is able to fluoroscopy and radiography from the front and the side
②　The system should be able to safely diagnose and treat colonic lung hypertonia thrombotic pulmonary artery
③　The system should have a function to examine the whole body, from head to peripheral, without moving the patient
④　The system should carry a flat-panel detector which has high X-ray conversion efficiency and density resolution
⑤　The flat-panel detector should be able to diagnose and treat by Right heart catheter
⑥　The digital image processing system should have the function to assist the Non-vascular treatment and vascular analysis function
⑦　The system should have a large LCD which is able to project multiple images
⑧　The system should be able to reconstruct by using 3D image data
⑨　The system should be able to connect with existing HIS, RIS, PACS, and diagnosis system
⑩　The bidder should provide peripheral equipment that is effective for managing machines and devices
⑪　The bidder should have an organization which provides quick user support and adequate maintenance of the system
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M 6, July, 2018
⑹　Contact point for the notice : Masataka Tsujimoto, Finance and Accounting Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext. 2982

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年5月22日
国立研究開発法人
国立循環器病研究センター
理事長　小川　久雄

◎調達機関番号　820　◎所在地番号　27
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　全身用コンピュータ断層撮影装置　3式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年1月末日
⑸　調達に必要とされる基本的な要求要件
①　循環器病のモデル医療や世界の先端に立つ高度先駆的医療に対応したデュアルエナジー対応のシステムであること。
②　多列型検出器を有し、その検出器の列数が128スライス以上であること。
③　短時間で広範囲の撮影が可能で、頭部や心臓を高速に撮影できるシステムであること。
④　患者および術者の放射線被ばくを低減する機能を有すること。
⑤　診断、治療、医学研究、教育に必要な最新アプリケーションを有すること。
⑥　患者の監視、管理ができる機能を有すること。
⑦　逐次近似画像再構成の機能を有すること。
⑧　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、読影システムおよび、医用画像管理システム（ＰＡＣＳ）と接続可能であること。
⑨　装置および機器の運営管理に有用な周辺機器を備えていること。
⑩　装置の維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料の提供期限　平成30年7月6日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台5―7―1　国立研究開発法人国立循環器病研究センター　財務経理部契約第一係　辻本　真孝　電話06―6833―5012（内線）2982
メールアドレス　Sounding@ml.ncvc.go.jp
3　導入説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年5月22日から平成30年7月6日まで。