c　脊椎外科用ナビゲーション専用インスツルメントを有していること。
d　術中画像撮影装置対応ソフトウェアを有していること。
e　術中画像撮影装置との連携が可能であること。
B　移動型Ｃアーム
a　フラットパネルを搭載していること。
b　透視及びＤＲ撮影ができること。
c　ＤＳＡ撮影ができること。
d　病院既設のＨＩＳ、ＲＩＳ及びＰＡＣＳに接続可能であること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年８月24日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒683―8504 米子市西町36―１ 鳥取大学医学部経理・調達課病院調達係　坂本千亜紀　電話0859―38―7087
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年７月２日から平成30年８月24日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年７月９日13時30分
⑵　開催場所　鳥取大学医学部附属病院打合室３
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Intraoperative imaging based navigation system 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　Intraoperative imaging based navigation system
a　To be equipped with Spine navigation software.
b　To be equipped with ＥＮＴ navigation software.
c　To be equipped with Spine navigation instruments.
d　To be equipped with data handling software for intraoperative images.
e　To be capable of integration with an intraoperative imaging devices.
B　Mobile Ｃ-arm system with Flat panel detector
a　The Xray detector shall be flat panel detector.
b　The function of fluoroscopy and digital radiography are available.
c　The function of digital subtraction angiography is available.
d　To be possible to connect to existing ＨＩＳ,ＲＩＳand ＰＡＣＳ.
⑸　Time limit for the submission of the requested material : 17 : 00　24 August, 2018
⑹　Contact point for the notice : Chiaki Sakamoto, Hospital Procurement Section, Accounting and Procurement Division, Faculty of Medicine, Tottori University, 36―1 Nishi-cho Yonago-shi 683―8504 Japan, TEL 0859―38―7087

○第３号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　外部放射線照射システム　一式
⑶　調達方法　借入
⑷　導入予定時期　平成30年度12月以降
⑸　調達に必要とされる基本的な要求要件
A　放射線治療装置は、デュアルエネルギー仕様を有し、全身の腫瘍に対して高精度放射線治療（強度変調放射線治療、体幹部定位放射線治療、画像誘導放射線治療等）が可能な医療用加速装置であること。
B　遠隔支援が可能な３次元放射線治療計画システムを有すること。
C　画像誘導放射線治療システムは，独立した２つのＸ線撮影装置で撮影した画像により位置照合が可能なシステムであること。
D　治療計画用コンピュータ断層撮影装置は１スキャンで16スライス以上撮影することができる機能を有すること。
E　呼吸同期システムを有すること。
F　高精度放射線治療システムのＱＡ・ＱＣに必要な周辺機器を有していること。
G　既存の画像および情報ネットワークシステムと接続可能であり、情報の受け渡しができること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年８月24日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒683―8504 米子市西町36―１ 鳥取大学医学部経理・調達課病院調達係　坂本千亜紀　電話0859―38―7087
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年７月２日から平成30年８月24日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年７月９日14時00分
⑵　開催場所　鳥取大学医学部附属病院打合室３
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : External Radiation Therapy System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　The radiation therapy equipment can emit two or more different energy. And it can perform high precision radiotherapy (Intensity modulated radiation therapy, Stereotactic Body Radiation Therapy, Image-guided radiation therapy) for all kinds of brain and body tumors.
B　The system must have a 3D treatment planning system that can be remotely supported.
C　The image-guided radiation therapy system can provide target position matching by orthogonal beam obtained from two indipendent X-ray apparatus.
D　The Conputed Tomography system for treatment planning must have multi-detector row more than 16 slices.
E　The system must have a respiratory gating or synchronizing apparatus.
F　The system must have the necessary related peripheral devices for ＱＡ／ＱＣ of high precision radiotherapy.
G　The system must be able to communicate online with the hospital／radiology information system.
⑸　Time limit for the submission of the requested material : 17 : 00　24 August, 2018
⑹　Contact point for the notice : Chiaki Sakamoto, Hospital Procurement Section, Accounting and Procurement Division, Faculty of Medicine, Tottori University, 36―1 Nishi-cho Yonago-shi 683―8504 Japan, ＴＥＬ 0859―38―7087

資料提供招請に関する公表

西日本電信電話株式会社公告第８号

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

なお、本招請は資料等提供者の機密情報及び財産的情報の提出を求めるものではありません。

平成30年７月２日

西日本電信電話株式会社
大阪病院長　前田　恵治

◎調達機関番号　126　◎所在地番号　27

○整理番号　西人経大病第128号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品名称及び数量　ＰＥＴ―ＣＴ装置　一式
⑶　調達方法　購入
⑷　導入予定時期　平成31年３月
⑸　調達に必要とされる基本的な要求用件
（Ａ１）　ＰＥＴ―ＣＴ装置