M　４Ｋ３Ｄシステムのカメラは、顕微鏡に内蔵していること。
N　３Ｄの表示可能なフォーマットは、汎用性の高いSide by Side方式であること。
O　顕微鏡は，本院が指定するオペレーションチェアに接続させること。
P　顕微鏡は，オペレーションチェアのフットスイッチ経由で操作ができる機能を有していること。
Q　顕微鏡本体にあるコンセントに接続するだけで迅速に血管の観察が行える硬性鏡システムを有していること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和２年６月８日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒770―8503　徳島市蔵本町２―50―１　国立大学法人徳島大学病院経理調達課調達係　梶浦　智子　電話088―633―9559

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和２年４月21日から令和２年６月８日まで。
⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和２年４月27日15時00分
⑵　開催場所　国立大学法人徳島大学病院入札室・会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Surgical microscope equipped with 4K3D visualization system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must provide all apochromatic optics.
B　The system must provide motorized magnification adjustment with zoom ratio 1 : 6 controlled by either handgrips or footswitch.
C　The system must provide variable adjustment of the working distance from 200mm to 625mm.
D　The system must provide autofocus function with laser measurement method within the accuracy of ±0.5mm.
E　The system must provide a fully integrated 4K3D camera as standard configuration.
F　The microscopes' illumination system must provide an additional light beam path to brighten up shadowed areas in the field of view.
G　When tilting the microscope laterally the co-observer tube must not move.
H　The system must complete the automatic balancing of all system axes at the push of a button on HD graphic touch screen.
I　Draping of the system must be facilitated by an automatic air vacuum.
J　The system must provide a fully integrated active vibration control system.
K　The system must offer a fully integrated intraoperative fluorescence guided tumor resection module.
L　The system must offer a fully integrated intraoperative fluorescence-based video angiography module.
M　The camera part of the 4K3D visualization attachable to the surgical microscope.
N　The 3D monitor must display in side-by-side format　recognized by the most 3D monitors.
O　This microscope can connect to the operationchair our hosupital specifiled.
P　This microscope has functions that can be operated via footswich of operationchair.
Q　This microscope has Regid endoscopesystem. This system can observe blood vessels quickly just plug in the outlet of microscope.
⑸　Time limit for the submission of the requested material : 17 : 00　8 June, 2020
⑹　Contact point for the notice : Tomoko Kajiura, Supplies, University Hospital, Accounting and Supply Affairs Section , National University Corporation, Tokushima University, 2―50―1 Kuramoto-cho Tokushima-shi 770―8503 Japan, TEL 088―633―9559

## 意見招請に関する公示

次のとおり調達に関する仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和２年４月21日
国土交通省航空局航空ネットワーク部
国際航空課長　大沼　俊之
◎調達機関番号　020　◎所在地番号　13
**○国空業第５号**
1　調達内容
⑴　品目分類番号　14、71、27
⑵　調達等件名及び数量
航空機運航に係る申請等の電子化システムの製造及び調整
2　意見の提出方法
⑴　意見の提出期限　令和２年５月19日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒100―8918 東京都千代田区霞が関２－１－３　国土交通省航空局予算・管財室契約係　田場　吉人　電話03―5253―8111内線48655
3　仕様書案の送付
⑴　交付期間　令和２年４月20日から令和２年５月18日まで
⑵　交付場所　上記２⑵と同じ
4　Summary
⑴　Classification of the product to be procured : 14, 71, 27
⑵　Nature and quantity of the products to be purchased : Management of Information for Koku system, 1 set
⑶　Time-limit for the Submission of Comments : 17 : 00 May 19, 2020
⑷　Contact point for the notice : Yoshito Taba Contract section, Budget and Property Administration Office, Civil Aviation Bureau, Ministry of Land, Infrastructure, Transport and Tourism 2―1―3 Kasumigaseki, Chiyoda-ku, Tokyo 100―8918 Tel : ＋81―3―5253―8111 (Ext. 48655)

## 意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和２年４月21日
国立大学法人群馬大学事務局長　児島　昌樹
◎調達機関番号　415　◎所在地番号　10
**○第１号**
1　調達内容
⑴　品目分類番号　22、32
⑵　借入物品及び数量
血管造影システム　一式
2　意見の提出方法
⑴　意見の提出期限　令和２年５月22日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒371―8510 前橋市荒牧町４丁目２番地　群馬大学財務部経理課専門職員　関谷　伸次　電話027―220―7609
3　仕様書案の交付
⑴　交付期間　令和２年４月21日から令和２年５月22日まで。
⑵　交付場所　上記２⑵に同じ。
4　仕様書案の説明会
⑴　開催日時　令和２年４月24日11時00分
⑵　開催場所　群馬大学医学部附属病院　病院中会議室
5　Summary
⑴　Classification of the products to be procured : 22, 32
⑵　Nature and quantity of the products to be rent : angiography system 1 Set
⑶　Time limit for the submission of comments : 17 : 00　22 May, 2020
⑷　Contact point for the notice : Shinji SEKIYA, Accounting Division, Gunma University, 4―2 Aramaki-machi Maebashi-shi 371―8510 Japan, TEL 027―220―7609