⑶　調達方法　購入等
⑷　導入予定時期　令和３年度10月以降
⑸　調達に必要とされる基本的な要求要件
A　検体搬送システムは、検体検査管理システムからの情報に基づいて、検体に前処理を施し、各測定装置へ自動搬送ができること。
B　必要な分析装置を搬送ラインに接続し、効率化を図った構築ができること。
C　検体検査システムとして、生化学・免疫・血液・一般検査に関する検体検査の自動測定ができること。
D　時間外検査に用いる分析装置は24時間稼働が可能なこと。
E　既存の病院情報システムとの連携ができること。
F　検体検査管理システムとして、検体搬送システムや検体検査システムから発生する情報の総合的な管理ができること。
G　患者の情報保護のため、セキュリティ対策及びシステムダウン対策が施されていること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和２年10月12日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課　林　寿美子　電話075―751―3024
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和２年８月26日から令和２年10月12日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和２年９月３日15時00分
⑵　開催場所　京都大学医学部附属病院会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Integrated Clinical Laboratory system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Automated transport system should make the preparation of samples and transport them to analytical devices automatically, according to the information from Laboratory management system.
B　To be able to construct the Sample Transportation System connected to analytical device to work effectively.
C　Laboratory analytic system should perform biochemical, immunological, hematological and general examinations, automatically.
D　The analysis equipment used after business time should have 24-hour operation ability.
E　Integrated Clinical Laboratory system should be online-connected to Kyoto University Hospital Information system .
F　Laboratory management system should integratedly manage the information from Automated transport system and Laboratory Analytic system.
G　Extreme care should be paid to the security of patient information and the prevention against the system crashing.
⑸　Time limit for the submission of the requested material : 17 : 00　12 October, 2020
⑹　Contact point for the notice : Sumiko Hayashi, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3024

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和２年８月26日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第６号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
手術室システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和３年度10月以降
⑸　調達に必要とされる基本的な要求要件
A　無影灯、シーリングペンダント、術野映像システム、手術室環境支援システム、術中動画記録システム、手術台、保温保冷庫で構成されること。
B　無影灯は、術野の高い無影性、色の再現性、放射熱の軽減の要件を満たす装置であること。
C　術野映像システム及び術中動画記録システムは、近年の映像技術の進歩に見合った、映像データ利用の利便性が図れるシステムであること。また、単なる手術の記録のみならず、医療安全や教育に使用するために適した性能・機能を有すること。
D　術中動画記録システムは、術野映像や手術室内監視モニター映像から器材やスタッフの導線を解析し、体内遺残の有無や業務の効率性、術中環境の最適化を検討するために適した性能・機能を有すること。
E　手術室環境支援システムは、手術室内の清浄度等を常時監視して、手術に至適な環境を維持するためのシステムであること。
F　安全性が保障されているものであること。
G　最新の性能・機能を有するものであること。
H　狭い部屋を有効活用し、手術が安全で効率良く行えるように、手術室内に配置できるものであること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和２年10月12日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課　林　寿美子　電話075―751―3024
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和２年８月26日から令和２年10月12日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和２年９月３日15時30分
⑵　開催場所　京都大学医学部附属病院会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Operating room system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Be comprised of an operating light, a sealing pendant, an operation field picture system, operating room environment support system, a perioperative video record system, an operating table, a thermal insulation cooling box.
B　An operating light being a device meeting the high no shadow characteristics of the operation field, color reproducibility, requirements of the reduction of the radiant heat.
C　An operation field picture system and the perioperative video record system being the systems which convenience of the use of picture data appropriate to the progress of the recent imaging technique can plan. Also, as well as a record of the simple surgery, have suitable performance, function to use it for medical security and education.