(4) Basic requirements of the procurement :
A The system (blood pump, inflow cannula, outflow graft, controlling unit included) is designed to be used to patients who are at risk of imminent death from severe left ventricular failure with limited drug effect, who qualify for heart transplant and have no other option but heart transplant for their survival.
B The blood pump is continuous flow centrifugal type, which constitutes of rotating impeller with lubricating liquid "Fluid Circulation System" around the hydrodynamic bearing.
C It is intended to meet the following configurations.
a Blood Pump
b Inflow Cannula
c Outflow Graft
d Controller
e Controller Connection Kit
f Battery
g Emergency Battery
h Emergency Controller
i AC／DC Adapter
j Battery Charger
k Coring knife
(5) Time limit for the submission of the requested material : 17 : 00　27 October, 2020
(6) Contact point for the notice : Takashi Yamaguchi, Procurement Section, Accounting Division, The University of Tokyo Hospital, 7—3—1 Hongo Bunkyo-ku Tokyo 113—8655 Japan, TEL 03—5800—8617

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和２年９月11日
国立大学法人鳥取大学長　中島　廣光
◎調達機関番号　415　◎所在地番号　31
○第１号
１　調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　高精度放射線治療システム　一式
(3) 調達方法　借入
(4) 導入予定時期　令和３年度９月以降
(5) 調達に必要とされる基本的な要求要件
A 放射線治療装置は、高精度放射線治療（強度変調放射線治療、定位放射線治療、画像誘導放射線治療等）が可能な医療用加速装置であること。
B 遠隔支援が可能な３次元放射線治療計画システムを有すること。
C 画像誘導放射線治療システムは，体表面、骨構造、腫瘍の位置情報による位置照合が可能なシステムであること。
D 治療計画用コンピュータ断層撮影装置は１スキャンで16スライス以上撮影することができる機能を有すること。
E 呼吸同期システムを有すること。
F 多発脳腫瘍に対する脳定位放射線治療が可能なシステムであること。
G 高精度放射線治療システムのＱＡ・ＱＣに必要な周辺機器を有していること。
H 既存の画像および情報ネットワークシステムと接続可能であり、情報の受け渡しができること。
２　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　令和２年10月27日17時00分（郵送の場合は必着のこと。）
(2) 提供先　〒683—8504 米子市西町36—１ 鳥取大学米子地区事務部経理・調達課病院調達係　紀村　昌弘　電話0859—38—7087
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　令和２年９月11日から令和２年10月27日まで。
(2) 交付場所　上記２(2)に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
(1) 開催日時　令和２年９月17日14時00分
(2) 開催場所　鳥取大学医学部附属病院会議室
(5)
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be rent : High Precision Radiotherapy System 1 Set
(3) Type of the procurement : Rent
(4) Basic requirements of the procurement :
A The linear accelerator should be possible to perform high precision radiotherapy (Intensity modulated radiation therapy, Stereotactic Radiation Therapy, Image-guided radiation therapy).
B The system must have a 3D treatment planning system that can be remotely supported.
C The image-guided radiation therapy system should be possible to collate based on body surface, bone structure, and tumor position information.
D The Computed Tomography system for treatment planning must have multi-detector row more than 16 slices.
E The system must have a respiratory gating or synchronizing apparatus.
F The system should be possible to perform a brain stereotactic radiation therapy for multiple brain tumors.
G The system must have the necessary related peripheral devices for QA／QC of high precision radiotherapy.
H The system must be able to communicate online with the hospital／radiology information system.
(5) Time limit for the submission of the requested material : 17 : 00　27 October, 2020
(6) Contact point for the notice : Masahiro Kimura, Hospital Procurement Section, Accounting and Procurement Division, Yonago Administration Office, Tottori University, 36—1 Nishi-cho Yonago-shi 683—8504 Japan, TEL 0859—38—7087

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和２年９月11日
大学共同利用機関法人
高エネルギー加速器研究機構長　山内　正則
◎調達機関番号　416　◎所在地番号　08
○第３号
１　調達内容
(1) 品目分類番号　71、27
(2) 購入等物品及び数量　計算機等の運用管理支援及び監視業務　一式
２　意見の提出方法
(1) 意見の提出期限　令和２年10月２日17時00分（郵送の場合は必着のこと。）
(2) 提出先　〒305—0801 つくば市大穂１－１ 大学共同利用機関法人高エネルギー加速器研究機構財務部契約課契約第五係　板山　裕之　電話029—864—5148
３　仕様書案の交付
(1) 交付期間　令和２年９月11日から令和２年10月２日まで。
(2) 交付場所　上記２(2)に同じ。
４　仕様書案の説明会
(1) 開催日時　令和２年９月16日14時30分
(2) 開催場所　大学共同利用機関法人高エネルギー加速器研究機構管理棟財務部入札室
５　Summary
(1) Classification of the services to be procured : 71, 27
(2) Nature and quantity of the services to be required : Operation and Remote monitoring Service for Computer Environment 1 Set
(3) Time limit for the submission of comments : 17 : 00　2 October, 2020
(4) Contact point for the notice : ITAYAMA Hiroyuki, Contracts Unit 5, Contracts Division, Financial Department, Inter-University Research Institute Corporation High Energy Accelerator Research Organization, 1—1 Oho Tsukuba-shi 305—0801 Japan, TEL 029—864—5148