③ Serial Angiographic System 28 sets
④ Gamma Camera System 15 sets
⑤ High Radiation Therapy System 6 sets
⑥ X-ray fluoroscope Radiography System 30 sets
⑦ General Radiographic System 24 sets
⑧ Digital X rays equipment to mammography 16 sets
⑨ Mobile X-ray Surgical Imaging System 19 sets
⑶ Type of the procurement : Purchase
⑷ Basic requirements of the procuremen :
① Multi-slice Computed Tomography System
A The system should be capable of scanning more than 16 slices data during one rotation
B The system should have ability that could scan either whole lung or abdomen in few seconds
② Magnetic Resonance Imaging System
A Magnetic field strength should be 1.5 Tesla or more
B The images of examination in head, body and extremities are enable to examine speedily with high resolution images
C The console unit must provide high performance capability and exceptional in operation
③ Serial Angiographic System
A The system should be designed safely for interventional procedures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography
B C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary-direction
C The flat-panel detector should be especially designed to fulfill the interventional angiography
D The digital imaging processing system should have supporting and analysis function for intravascular treatment
④ Gamma Camera System
A The system should be multiple digital gamma camera
B The system should have function of SPECT and Whole body acquisition
C The system should equip the fusion software and quantitative application replete
⑤ High Radiation Therapy System
A The system should be capable to multipurpose treatment
B The system should equip multi-leaf collimator with high shielding performance
C The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiation therapy
⑥ X-ray fluoroscope Radiography System
A The system is available in fluoroscopic radiography
B The system should be capable of recording dynamic images
C The system should be capable of positioning the bed position to horizontal
D The system should equip the flat-panel detector
⑦ General Radiographic System
A The system should equip X-ray tube
B The system should equip a table for standing and lying position, in addition, radiography of whole body could be able to perform speedily and safely
C The system should equip the flat-panel detector
⑧ Digital X rays equipment to mammography
A The system should be digital X rays equipment to mammography
B The system must be able to digital stereotactic mammography
C The system should equip the high resolution monitor system for mammography diagnosis
⑨ Mobile X-ray Surgical Imaging System : The system should be digital mobile X-ray surgical imaging system
⑸ Time-limit for the submission of the requested material : 5 : 00 P.M. March 1, 2021
⑹ Contact point for the notice : Yasuyuki Hotta, Chief Clerk, Supply Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2―5―21 Higashigaoka Meguro-ku Tokyo 152―8621, Japan. TEL 03―5712―5065

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和３年１月12日

国立大学法人金沢大学長　山崎　光悦

◎調達機関番号　415　◎所在地番号　17

○第５号

１　調達内容

⑴　品目分類番号　14

⑵　導入計画物品及び数量　電子情報教育用コンピュータシステム　一式

⑶　調達方法　借入

⑷　導入予定時期　令和４年度４月以降

⑸　調達に必要とされる基本的な要求要件

本システムは、学類及び大学院における情報教育を目的とし、コンピュータリテラシー教育、プログラミング演習、解析実験等に用いられるものであり、サーバ用計算機群、実習用計算機群、ネットワーク機器群及びプレゼンテーション用機器群からなる。

A　サーバ用機器群は、ファイルサーバ、アカウントサーバ、ネームサーバ、メールサーバ、ウェブサーバ等の機能を有し、1,000名以上の利用者に対する十分な資源、及び100名以上の利用者による一斉使用に対して十分に高速な応答を実現できる処理能力を有すること。また、利用者により持ち込まれる可搬型計算機等に仮想デスクトップ等のサービスを提供する機能も含むこと。

B　実習用計算機群は、100台以上の計算機からなり、高解像度を有するカラーディスプレイ、キーボード、マウスなどの入出力装置を備えていること。また、サーバ用計算機群が提供するサービスを利用するための機能を有すること。

C　ネットワーク機器群は、サーバ用計算機群、実習用計算機群を相互に接続するとともに、ファイアウォールを介して金沢大学既設ＬＡＮと接続し、高速なデータ通信を実現できる性能を有すること。また、可搬型計算機等の一時的な接続を提供するための機器を含むこと。

D　プレゼンテーション用機器群は、大型ディスプレイ、音響機器等からなり、実習用計算機や可搬型計算機等の映像出力を高輝度かつ高精細で表示可能であること。

E　故障、停電、セキュリティ等について対策がなされており、安定した運用が可能なシステムであること。また、システムの障害に対して迅速かつ万全な保守が可能であること。

F　システムの管理・運用に対して、省力化、省エネルギー化、省スペース化が図られていること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和３年２月15日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒920―1192 金沢市角間町　金沢大学理工系事務部会計課調達係長　松田　光弘　電話076―234―6846

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和３年１月12日から令和３年２月15日まで。

⑵　交付場所　上記２⑵に同じ。