⑥ The system should be able to set up to the auto-position in each protocol.
⑦ The high voltage of X-ray tubes should be generated by inverter systems.
⑧ The X-ray tube should have the function of a small focus and large current, and the maximum anode thermal storage capacity should be sufficient.
⑨ A countermeasure to minimize the X-ray exposure of patients, operators, and so on should be provided.
⑩ The system should be able to preserve images sufficiently. Also it must be linked to Radiolody Information Systems-Picture Archiving and Communication Systems server of the JSDF Yokosuka hospital.
⑪ The X-ray angiography device must be in a single plane.
⑫ The digital images output format from the system should be based on latest DICOM.

⑸ Time-limit for the submission of the requested material : 5 : 00 P.M. 13 May, 2022
⑹ Contact point for the notice : Hiroyuki Tawara, Procurement Officer, Medical Planning Div., Maritime Staff Office, Ministry of Defense, 5―1 Ichigayahonmuracho Shinjuku-ku Tokyo 162―8803, Japan. TEL 03―3268―3111 ext. 50081

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和４年３月29日

契約責任者
独立行政法人国立印刷局
財務担当理事　倉林　健二

◎調達機関番号　562　◎所在地番号　13

1　調達内容
⑴　品目分類番号　12
⑵　購入等件名及び数量
紙料調製設備制御部外改造　一式

2　意見の提出方法
⑴　意見の提出期限　令和４年４月22日午後４時（郵送の場合は必着のこと。）
⑵　提出先　〒105―8445 東京都港区虎ノ門二丁目２番５号　独立行政法人国立印刷局財務部契約課施設契約係担当　宮代　英輔　電話03―6893―4619

3　仕様書案の交付
⑴　交付期間　令和４年３月29日から令和４年４月19日まで。
⑵　交付場所　上記２⑵に同じ

4　仕様書案の説明会
⑴　開催日時　令和４年４月21日午前10時
⑵　開催場所　独立行政法人国立印刷局財務部入札室

5　Summary
⑴　Classification of the products to be procured : 12
⑵　Nature and quantity of the products to be purchased : Paper material preparation equipment control section and other remodeling : 1 set
⑶　Time-limit for the submission of comments : 4 : 00 P.M. 22 April 2022
⑷　Contact point for the notice : Eisuke Miyashiro, Procurement Contract Division, Financial Affairs Department, National Printing Bureau, Incorporated Administrative Agency, 2―2―5 Toranomon, Minato-ku, Tokyo 105―8445 Japan TEL 03―6893―4619

意見招請に関する公示

次のとおりコンピュータ製品及びサービスの調達に係る仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和４年３月29日

国立研究開発法人宇宙航空研究開発機構
第一宇宙技術部門
ＧＯＳＡＴ―ＧＷプロジェクトチーム
プロジェクトマネージャ　小島　　寧

◎調達機関番号　587　◎所在地番号　13

○第 21―042 号

1　調達内容
⑴　品目分類番号　71、27
⑵　調達物品及び数量　温室効果ガス・水循環観測技術衛星（ＧＯＳＡＴ―ＧＷ）　衛星管制システム用計算機の賃貸借・保守・据付（一式）

2　意見の提出方法
⑴　意見の提出期限　令和４年４月19日15時00分
⑵　提出先　〒305―8505 茨城県つくば市千現２丁目１番１　国立研究開発法人宇宙航空研究開発機構　筑波宇宙センター　第一宇宙技術部門　ＧＯＳＡＴ―ＧＷプロジェクトチーム　（正）山本　泰久　（副）宮本　裕行　電話（正）070―3117―4694　（副）070―1170―2533　ＦＡＸ029―868―5987
電子メール
（正）yamamoto.yasuhisa@jaxa.jp
（副）miyamoto.hiroyuki@jaxa.jp

3　仕様書案の交付
⑴　交付期間　令和４年３月29日から令和４年４月19日まで
⑵　交付場所　上記２⑵に同じ

4　仕様書案の説明会
⑴　開催日時　令和４年４月４日10時00分（要事前申込）（参加任意）
⑵　開催場所　筑波宇宙センター　総合開発推進棟２Ｆ　小会議室３

5　Summary
⑴　Classification of the products to be procured : 71, 27
⑵　Nature of the products to be procured : Leasing, maintenance, and installation of computers for GOSAT―GW satellite control systems
⑶　Time limit for the submission of comments : 15 : 00 (JST) April 19, 2022
⑷　Contact Point for the notice : Yasuhisa Yamamoto, Hiroyuki Miyamoto GOSAT―GW Project Team Space Technology Directorate I, Japan Aerospace Exploration Agency, Tsukuba Space Center (TKSC) 2―1―1, Sengen, Tsukuba, Ibaraki, 305―8505 Japan TEL ＋81―70―3117―4694 ＋81―70―1170―2533 FAX ＋81―29―868―5987
Email yamamoto.yasuhisa@jaxa.jp
miyamoto.hiroyuki@jaxa.jp

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和４年３月29日

独立行政法人国立病院機構
九州医療センター院長　森田　茂樹

◎調達機関番号　597　◎所在地番号　40

○第 12808 号（№12808）

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　消化器内視鏡症例単価払い賃貸借契約　一式

2　意見の提出方法
⑴　意見の提出期限　令和４年４月27日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒810―8563 福岡県福岡市中央区地行浜１丁目８番１号　独立行政法人国立病院機構九州医療センター企画課長　丸山　誠二　電話092―852―0700　内線1007

3　仕様書案の交付
⑴　交付期間　令和４年３月29日から令和４年４月27日まで。
⑵　交付場所　上記２⑵に同じ。

4　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be leased : Gastrointestinal endoscopy case unit price payment lease contract 1 set
⑶　Time-limit for the submission of comments : 5 : 00 P.M. April 27, 2022
⑷　Contact point for the notice : Seiji Maruyama, Director, Accounting Division, National Hospital Organization Kyusyu Medical Center, 1―8―1 Jigyohama Chuo-ku Fukuoka-shi Fukuoka-ken 810―8563, Japan. TEL 092―852―0700 ext. 1007