b　遠心ポンプ
ⅰ　遠心ポンプはシステムベースより電源供給されること。
ⅱ　自己診断機能を有していること。
B　冷温水槽
人工心肺用水槽が独立していること。
C　体外循環用血液ガス分析装置
血液ガス、電解質、酸素飽和度、ヘモグロビン、ヘマトクリット値をオンラインで連続測定できること。
D　人工心肺自動記録装置
人工心肺データ、患者生体情報を自動取り込みできること。
E　頭蓋内酸素飽和度モニタ
ⅰ　４ch以上測定できること。
ⅱ　成人から新生児まで測定できること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和４年５月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒036―8563 青森県弘前市本町53 弘前大学医学部附属病院経理調達課　役務担当　工藤かおり　電話0172―39―5188

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和４年４月11日から令和４年５月27日まで。
⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和４年４月15日11時30分
⑵　開催場所　弘前大学医学部附属病院小会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Cardiopulmonary Bypass System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　Cardiopulmonary bypass equipment
It should be configured roller pump and a centrifugal pump for selective cerebral perfusion and cardioplegia.
a　Roller pump
i　Separate pump head should be consist of large pump and small pump.
ii　It must have a self-diagnostic function.
b　Centrifugal pump
i　It should be supplied power from the system base.
ii　It must have a self-diagnostic function.
B　Heater-Cooler System
It must be independent from the system.
C　Blood Parameter Monitoring System circulations :
It must have a function that be measured continuously online blood gas, electrolytes, oxygen saturation, hemoglobin, hematocrit value.
D　Automatic recording device cardiopulmonary bypass.
It must have a function to uptake automatically the cardiopulmonary bypass date and the patient monitoring system date.
E　Intracranial oxygen saturation monitor
i　It must be able to measure more than 4ch.
ii　It must be able to monitor one from adult to neonatal.
⑸　Time limit for the submission of the requested material : 17 : 00　27 May, 2022
⑹　Contact point for the notice : Kaori Kudo, Accounting Procurement Division, Hirosaki University Hospital, 53 Hon-cho Hirosaki-shi Aomori Japan 036―8563 Japan, TEL 0172―39―5188

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和４年４月11日
国立大学法人群馬大学
医学部附属病院長　齋藤　　繁
◎調達機関番号　415　◎所在地番号　10
○第２号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　バイプレーン血管撮影システム　一式
⑶　調達方法　購入等または借入
⑷　導入予定時期　令和５年度４月以降
⑸　調達に必要とされる基本的な要求要件
A　血管撮影装置は床置型正面Ｃアーム支持装置及び天井懸垂型側面Ｃアーム支持装置を組み合わせたバイプレーン方式であり、大視野フラットパネルディテクタ（ＦＰＤ）及びＸ線管装置を搭載すること。
B　術者用の大型モニタを備えており、各種の映像入力を一元表示できるシステムであること。
C　高度な検査及び治療に対応するため、最新の３Ｄ撮影機能や２Ｄ／３Ｄアプリケーションを有していること。
D　既存の情報システム（ＲＩＳ／ＰＡＣＳ）と接続し、情報通信を行うこと。
E　大容量のデータ保存が可能であること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和４年５月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒371―8511 前橋市昭和町３丁目39―22　群馬大学昭和地区事務部管理運営課用度係　内田　聡美　電話027―220―7748

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和４年４月11日から令和４年５月27日まで。
⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和４年４月21日13時00分
⑵　開催場所　オンライン開催

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be required : Biplane angiography system 1 Set
⑶　Type of the procurement : Rent or Purchase
⑷　Basic requirements of the procurement :
A　Biplane angiography system that has a floor-stand frontal C-arm system and a ceiling-suspended lateral C-arm system, and it must be equipped with large field flat panel detectors (FPD) and X-ray tubes.
B　A large display for operators must be equipped and can display multiple inputs simultaneously.
C　Angiography must be equipped latest 3D image acquisition as well as 2D／3D applications to enable advanced treatment and examination.
D　The angiography system must be connected to the existing information systems (RIS／PACS).
E　The angiography system should have a large data storage space available.
⑸　Time limit for the submission of the requested material : 17 : 00　27 May, 2022
⑹　Contact point for the notice : Satomi UCHIDA, Procurement Section, Gunma University, 3―39―22 Showa-machi Maebashi-shi 371―8511 Japan, TEL 027―220―7748

第明治三十五種郵便物認可