c　This system should be based on the DICOM standard 2014 and the Integration Profile which is defined by IHE—J (Integrating the Healthcare Enterprise Japan).

d　This system must be durable for consecutive utilization with a reliable and appropriate performance on the principle of high availability.

⑸　Time limit for the submission of the requested material : 17 : 00　11 July, 2022

⑹　Contact point for the notice : Shigehiro Akiyama, Procurement Services Office, Finance Department, Tohoku University, 2—1—1　Katahira　Aoba-ku　Sendai-shi 980—8577 Japan, TEL 022—217—4869

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和４年５月26日

国立大学法人山口大学長　谷澤　幸生

◎調達機関番号　415　◎所在地番号　35

○第１号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　自動採血管準備システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和５年３月

⑸　調達に必要とされる基本的な要求要件

Ａ．職員や患者自身が血液や尿検査の受付をできるシステムであること

Ｂ．採血情報に基づき、22種以上の採血管にバーコードラベルを自動貼付できるシステムであること

Ｃ．システムは連結した10台以上の採血台システムで構成されていること

Ｄ．病棟分の採血管準備は、１患者ごとにビニール袋へ収納できるシステムであること

Ｅ．整理券発行機能を有し、受付順に連続番号での採血整理券を発行することができること

Ｆ．42インチ以上のカラー液晶モニター３台以上で構成された外待ち表示案内システムで患者待合誘導ができること

Ｇ．採血採尿受付機（２台以上）を有し、患者ＩＤカード（受診カード）もしくはバーコードを読み取ることで、採血整理券やバーコード付き尿カップを自動発行できること

Ｈ．検体情報統括管理システム機能を有し、自動採血管準備装置から発行されたラベル内部のＩＣチップを一括読取できること

Ｉ．尿検体情報統括管理システム機能を有し、尿カップのバーコードを読み取り、到着確認が可能であること

Ｊ．有人受付での採血管および尿カップの発行が可能であること

Ｋ．採血人数や採血待ち時間などのデータを記録し、抽出することが可能であること

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和４年７月11日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒755—8505 山口県宇部市南小串１－１－１　国立大学法人山口大学医学部管理運営課契約管理係　竹原　芳維　電話 0836—22—2031

E-mail　me215@yamaguchi-u.ac.jp

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和４年５月26日から令和４年７月11日まで。

⑵　交付場所　上記２⑵に同じ。

４　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

５　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Automatic blood collection tube preparation system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A. The system should have a capability that staffs or patients by themselves can register blood and urine tests.

B. The system can sticks barcode labels on 22 or more kinds of tubes automatically based on test order.

C. The system shall consist of 10 or more connected blood collection table systems.

D. The preparation of blood collection tubes for the ward should be a system that can be stored in a plastic bag for each patient.

E. It has a numbered ticket issuing function and can issue blood sampling numbered tickets with serial numbers in the order of reception.

F. Patient waiting guidance can be provided by the outside waiting display guidance system consisting of three or more 42-inch or larger color LCD monitors.

G. Blood collection and urine collection reception machines (2 or more units), and by reading the patient ID card (consultation card) or barcode, it is possible automatically issue blood collection numbered tickets and urine cups with barcodes.

H. It has a sample information management system function and can collectively read the IC chips inside the label issued by the automatic blood collection tube preparation device.

I. It has a urine sample information management system function, and it is possible to read the barcode of the urine cup and confirm the arrival.

J. It is possible to issue blood collection tubes and urine cups at a staffed reception.

K. It is possible to record and extract data such as the number of patients to collect blood and the waiting time for blood collection.

⑸　Time limit for the submission of the requested material : 17 : 00　11 July, 2022

⑹　Contact point for the notice : Yoshimasa Takehara, Procurement, Accounting Division, School of Medicine, Yamaguchi University, 1—1—1 Minamikogushi Ube-shi 755—8505 Japan, TEL 0836—22—2031 E-mail me215@yamaguchi-u.ac.jp

○第２号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　生化学自動分析装置　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和５年３月

⑸　調達に必要とされる基本的な要求要件

①　システムの構成

Ａ．生化学自動分析装置

Ｂ．電解質自動分析装置（生化学自動分析装置内蔵型ユニット）

Ｃ．無停電電源装置（生化学自動分析装置本体用）

②　システムの性能

Ａ．当院が行う生化学検査を遅延なく処理する検査処理能力を有すること

Ｂ．必要に応じて、血清、血漿、尿、髄液等の材料において分析が可能であること

Ｃ．既設検査システムと接続できること

Ｄ．外部搬送ラインと直接接続できること

Ｅ．24時間稼働に対応できる装置であること

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和４年７月11日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒755—8505 山口県宇部市南小串１－１－１　国立大学法人山口大学医学部管理運営課契約管理係　竹原　芳維　電話 0836—22—2031

E-mail　me215@yamaguchi-u.ac.jp

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和４年５月26日から令和４年７月11日まで。

⑵　交付場所　上記２⑵に同じ。

４　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

５　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Clinical Chemistry Analyzer 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

①　System Configuration

A. Clinical Chemistry Analyzer

B. Automatic electrolyte analyzer (built-in unit for a Clinical Chemistry Analyzer)