タッフの視認にも十分なサイズの大型液晶モニターで、鮮明な解像度を有するＨＤ（High Definition）、および、３Ｄ（Three Dimensions）信号の画像も表示可能なものとする。

b　外部モニターは、ＳＤＩ（３Ｇ対応）、ＤＶＩ－Ｄ、ＨＤＭＩの端子を有していること。

c　外部録画装置は、内蔵とＵＳＢ接続のハードディスクドライブ、内蔵とＵＳＢ接続の光学式ドライブ、ＵＳＢメモリーの記憶媒体に録画可能で、２つの記憶媒体に同時録画可能な機能を有するものとする。なお、録画装置は、ＵＳＢ接続のタッチモニターで操作可能なものとする。

C　納品後は、適切な保守・点検を提供する体制を有すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和４年８月16日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒113―8655 東京都文京区本郷７－３－１　東京大学医学部附属病院事務部管理課経理チーム　石橋　彰　電話03―5800―8617

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和４年７月１日から令和４年８月16日まで。

⑵　交付場所　上記２⑵に同じ。

4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Surgical Microscope System 2 Sets

⑶　Type of the procurement : purchase

⑷　Basic requirements of the procurement :

A　Surgical Microscope System 2 sets. The illumination of the light source installed in the Surgical microscope system is sent to the surgical field, and the illuminated surgical field is magnified by the objective lens, the variable magnification mechanism, and the eyepiece lens for observation by the operator, or the image captured by the built-in camera is projected on external monitor for digital observation instead of the eyepiece lens, it consists of a microscope base with a lens and camera, a control unit with a light source and control mechanism, and an arm that suspends the microscope base from the control unit, and shall have the following functions.

a　The control unit should be placed behind the operator, and the arm should be a mechanism to suspend the microscope base over the operator's head.

b　The handgrip type controller equipped on the microscope base and the wireless type foot switch controller can operate basic programs such as zoom and focus.

c　The microscope base shall have a built-in 3D (Three Dimensions) HD (High Definition) camera, and both sides of the control unit shall have a touch screen to project the captured images and a built-in recording function for captured images.

d　Have a removable binocular side speculum that allows simultaneous observation by non-operators.

e　Two high-power light source devices shall be mounted as illumination for the surgical field and shall have a function to automatically adjust brightness in conjunction with the working distance to the surgical field.

B　External Image System 2 sets. Connected to the Surgical Microscope System, it consists of an external monitor that projected the surgical field under observation, an external recording device that records the images, and a touch monitor for operating the external recording device, and shall have the following functions.

a　The external monitor is placed in the operating room and connected to the Surgical Microscope System, to project the surgical field under observation. The monitor shall be a large LCD monitor large enough to be viewed by the surgical staff at a distance and it shall also be capable of displaying clear resolution HD (High Definition) and 3D (Three Dimensions) signal images.

b　The external monitor must have SDI (3G compatible), DVI-D, and HDMI terminals.

c　The external recording device shall be capable of recording to a hard disk drive with built-in and USB connection, an optical drive with built-in and USB connection, and a USB memory storage medium, and shall have the ability to record to two storage media simultaneously. The recording device shall be operable with a USB-connected touch monitor.

C　After delivery, the supplier is required to have a system to provide adequate maintenance and inspection.

⑸　Time limit for the submission of the requested material : 17 : 00　16 August, 2022

⑹　Contact point for the notice :
ISHIBASHI Akira, Procurement Section, Accounting Division, The University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-ku Tokyo 113―8655 Japan, TEL 03―5800―8617

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和４年７月１日

国立大学法人　山梨大学長　島田　眞路

◎調達機関番号　415　◎所在地番号　19

○第１号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　注射薬自動払出システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和４年度３月

⑸　調達に必要とされる基本的な要求要件　詳細は導入説明書による。

2　資料及びコメントの提供方法　上記１⑵に物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和４年８月15日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒409―3898 山梨県中央市下河東1110番地　山梨大学医学域管理課　課長補佐　長濱　裕介　電話055―273―9204

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和４年７月１日から令和４年８月15日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和４年７月11日14時00分

⑵　開催場所　山梨大学医学部キャンパス内管理棟２階中会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Auto Ampoule Dispenser 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement : Refer to the instructions for detail

⑸　Time limit for the submission of the requested material : 17 : 00　15 August, 2022

⑹　Contact point for the notice : Yusuke Nagahama, Faculty of Medicine Accounting Division University of Yamanashi, 1110 Shimokato Chuo-City Yamanashi 409―3898 Japan, TEL 055―273―9204