⑽　外国における技術者資格をもって申請する場合には、別途国土交通省総合政策局建設市場整備課における建設コンサルタント業務等に関する国土交通大臣認定を受けた者のみを認めるものとする。

なお、参加表明書の提出期限までに当該認定を受けていない場合にも参加表明書を提出することができるが、この場合、参加表明書提出時に当該認定の申請書の写しを提出するものとし、当該業者が指名を受けるためには指名通知の日までに大臣認定を受け、認定書の写しを提出しなければならない。指名通知の日は令和４年10月17日（月）を予定する。

⑾　紙入札方式による参加を希望する場合の手続　紙入札方式による参加を希望する者は、上記４⑶①の期限までに、参加表明書等とともに紙入札方式参加（変更）届出書（電子入札留意事項様式１）を、上記４⑴に示す場所に持参又は郵送により提出しなければならない。

⑿　詳細は入札説明書による。

6　Summary

⑴　Official in charge of the contract of the procuring entity : Jun Nakamura Director General of Kyushu Branch, West Nippon Expressway Company Limited

⑵　Classification of the services to be procured : 42

⑶　Subject matter of the contract : Survey and Examination of Embankment Soil in kagoshima Expressway Office in Fiscal Year 2022

⑷　Time Limit for the submission of application forms and relevant documents for the qualification by electronic bidding system : 4 : 00 P.M. 28 September 2022 (if brought with you, 4 : 00 P.M. 28 September 2022. if by mail, 4 : 00 P.M. 28 September 2022)

⑸　Time Limit for the submission of proposals : 4 : 00 P.M. 27 October 2022

⑹　Time Limit for the submission of tenders by electronic bidding system : 11 : 00 A.M. 28 November 2022 (If brought with you, 11 : 00 A.M. 28 November 2022. If by mail 11 : 00 A.M. 28 November 2022)

⑺　The language used for application and inquiry shall be Japanese.

⑻　Contact point for tender documentation : Yasuaki Takamori, Assistant Manager of Accounting Division, General Affairs and Planning Department, Kyushu Branch, West Nippon Expressway Company Limited, 3―13―15, Hakataekihigashi, Hakata-Ku, Fukuoka City, Fukuoka Prefecture 812―0013 Japan Tel. 092―260―6115 Fax. 092―260―6118

招　　請

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和４年９月12日

国立大学法人岡山大学長　槇野　博史

◎調達機関番号　415　◎所在地番号　33

○第４号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　Ｘ線血管撮影システム　１式

⑶　調達方法　購入等

⑷　導入予定時期　令和５年度第１・四半期以降

⑸　調達に必要とされる基本的な要求要件

Ａ　平面検出器装置及びＸ線管装置を搭載したバイプレーン型血管撮影装置であること。

Ｂ　回転ＤＳＡの撮影及び３Ｄ構築が容易に行えること。

Ｃ　フラットパネル検出器・Ｘ線管部分に安全機能を有し、高速血管撮影が可能で高精細画像が得られること。

Ｄ　血管内治療を支援する機能及び解析機能を有すること。

Ｅ　患者及び医療従事者に対する被ばく低減機構を有すること。

Ｆ　面積線量計が装備されていること。

Ｇ　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、及び画像ネットワーク（ＰＡＣＳ）と接続すること。

Ｈ　線量レポート出力方式がＲＤＳＲに対応していること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和４年10月28日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒700―8558　岡山市北区鹿田町２－５－１　岡山大学病院経営・管理課調達グループ政府調達担当　中本　　誠　電話086―235―7540

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和４年９月12日から令和４年10月28日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和４年９月14日10時00分

⑵　開催場所　岡山大学医学部管理棟小会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : X-ray angiography System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The equipment should be a bi-plane angiography system with a flat panel detector and an X-ray tube.

B　The equipment must perform rotation DSA imaging and 3D construction efficiently.

C　The flat panel detector and X-ray tube must have safety functions. In addition, the equipment must be capable of high-speed angiography and high-definition images.

D　The equipment must have functions to support and analyze endovascular treatment.

E　The equipment must have the function of reducing X-rays exposure for patients and medical professionals.

F　The equipment must have an area dosimeter.

G　The equipment must connect our institution's HIS, RIS, and PACS systems.

H　The dose report output method must support RDSR.

⑸　Time limit for the submission of the requested material : 17 : 00　28 October, 2022

⑹　Contact point for the notice : NAKAMOTO Makoto, Procurement Section, Management Division, Okayama University Hospital, 2―5―1 Shikata-cho Kita-ku Okayama-shi 700―8558 Japan, TEL 086―235―7540

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和４年９月12日

国立大学法人山口大学長　谷澤　幸生

◎調達機関番号　415　◎所在地番号　35

○第３号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　脳神経外科手術用顕微鏡　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和５年11月

⑸　調達に必要とされる基本的な要求要件

Ａ．術中血管観察、術中腫瘍観察、術中蛍光血管撮影および血行動態解析ができる機能を有すること。

Ｂ．フルハイビジョンで高画質の記録ができること。

Ｃ．ナビゲーションリンクが可能であること。

Ｄ．オートバランス及びレーザー測光式オートフォーカスができる機能を有すること。