B　Thermal analyses should be possible.
C　This system is able to measure precise amounts of elements constituting organic matter.
D　This system is able to measure absorbance of liquid samples and absorption／reflection of solid samples, such as glasses, films, and so on.
E　This system is able to analyze molecular structure of a substance and obtain 2D and 3D images by detection of raman scattered light.
F　This system is able to analyze molecular structure of a substance at an atomic level based on the principle called nuclear magnetic resonance. Furthermore, reduction of liquid helium consumption is required.
G　A cylindrical container which contains strong electromagnetic coils immersed in liquid helium. Furthermore, reduction of liquid helium consumption is required.
⑸　Time limit for the submission of the requested material : 17 : 00　31 May, 2023
⑹　Contact point for the notice : Koichi TAKAHASHI, Procurement Section, Gunma University, 1―5―1 Tenjin-cho Kiryu-shi 376―8515 Japan, TEL 0277―30―1032

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和５年４月26日
国立大学法人岡山大学長　那須　保友
◎調達機関番号　415　◎所在地番号　33
○第１号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　高精度放射線治療システム　１式
⑶　調達方法　購入等
⑷　導入予定時期　令和５年度第３・四半期以降
⑸　調達に必要とされる基本的な要求要件
Ａ　強度変調放射線治療（ＩＭＲＴ）、定位放射線治療（ＳＲＴ）、体幹部定位放射線治療（ＳＢＲＴ）、画像誘導放射線治療（ＩＧＲＴ）、体表面画像誘導放射線治療（ＳＧＲＴ）による高精度放射線治療が可能であること。
Ｂ　４次元放射線治療による高精度放射線治療が可能であること。
Ｃ　あらゆる高精度放射線治療に対応する患者固定具及び補助具を備えること。
Ｄ　様々な高精度放射線治療が効率良く計画可能な放射線治療計画装置を備えること。
Ｅ　様々なモダリティの画像レジストレーション、スムーズな放射線治療計画レビュー、コンツーリング支援などの機能を有する放射線治療計画支援システムを備えること。
Ｆ　院内既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、医療用画像管理システム（ＰＡＣＳ）に互いに接続可能な放射線治療マネジメントシステムを備えること。
Ｇ　高精度放射線治療に必要な品質保証及び品質管理が実施可能な機器を備えること。
Ｈ　遠隔放射線治療計画に必要なシステムを備えること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和５年６月12日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒700―8558　岡山市北区鹿田町２―５―１　岡山大学病院経営・管理課調達グループ政府調達担当　中本めぐ美　電話086―235―7540
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和５年４月26日から令和５年６月12日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和５年４月28日10時30分
⑵　開催場所　岡山大学医学部管理棟小会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : High-Precise Radiation Therapy System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must have the capability to perform Intensity-Modulated Radiation Therapy (IMRT), Stereotactic Radiation Therapy (SRT), Stereotactic-Body Radiation Therapy (SBRT), Image-Guided Radiation Therapy (IGRT) and Surface Guided Radiation Therapy (SGRT).
B　The system must have the capability to perform high-precision 4-dimentional radiotherapy.
C　The system must have various kind of accessories and fixture device to perform high-precision radiotherapy.
D　The system must have radiation treatment planning system that create highly accurate radiation treatment plan quickly and efficiently.
E　The system must be equipped with a radiation therapy plan support system capable of image registration of various modality, smooth radiation therapy plan review, and support for contouring.
F　The system must be connected to the network of hospital information system (HIS), radiology information system (RIS) and picture archiving and communication system (PACS), and can exchange their data each other.
G　The system must have equipment capable of quality assurance and quality control for high-precise radiation therapy.
H　The system needs to have features for remote radiation therapy planning.
⑸　Time limit for the submission of the requested material : 17 : 00　12 June, 2023
⑹　Contact point for the notice : NAKAMOTO Megumi, Procurement Section, Management Division, Okayama University Hospital, 2―5―1 Shikata-cho Kita-ku Okayama-shi 700―8558 Japan, TEL 086―235―7540

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和５年４月26日
国立大学法人山口大学長　谷澤　幸生
◎調達機関番号　415　◎所在地番号　35
○第１号
１　調達内容
⑴　品目分類番号　24
⑵　導入計画物品及び数量　電界放出型走査電子顕微鏡　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和６年１月以降
⑸　調達に必要とされる基本的な要求要件
Ａ．電界放出型走査電子顕微鏡（FE-SEM）　一台
ａ　電子ビーム、電子レンズ、画像検出のための高性能な装置を有し、生物試料の高分解能による観察を可能とすること
ｂ　薄切切片の透過観察と高精細な画像取得を可能とすること
ｃ　生物試料を出来るだけ自然に近い状態で観察できることを可能とするような、低真空環境での高精細観察が可能なこと
Ｂ．エネルギー分散型X線元素分析装置　一台
ａ　生物試料の任意の部位における元素組成分析・元素マッピング分析を可能とすること
Ｃ．光／電子相関観察ユニット　一式
ａ　FE-SEMと共同獣医学部所有の倒立蛍光顕微鏡との間において、双方向での画像相関が可能なこと
ｂ　両機器の座標軸合わせを可能とし、観察画像の重ね合わせを可能とすること