⑨　装置の管理者、運用者に本件調達物品に対する技術指導を行うための研修に掛かる費用を含むこと。
⑩　当該仕様書に明記されていない事項であっても、本件調達品の稼動に必要な事項は怠りなく施行し、疑義については当センター担当者と協議し、指示を受けること。
⑪　機器運用開始１年間は、通常の使用により故障した場合の無償修理に応じること。
⑫　改正医療法（臨床検査室）に対する充分な配慮がなされている機器および試薬構成であること。
⑬　現行機器と操作方法が異なる場合、現行マニュアルと同様なマニュアル等の作成ならびに、日常業務・時間外業務に支障がないよう配慮すること。
⑭　今回調達の機器全てにおいて純水製造装置が必要な場合、装置１台あたりイオン交換樹脂を複数本使用した装置２台（同一メーカーの機器）に集約し、装置メンテナンスならびにトラブル時は相互に純水を使用可能な仕組みを準備すること。
⑮　調達機器については、原則無停電装置を準備すること。また、純水装置・制御端末も含め無停電装置を付け、全体で瞬間停電に対処できること。
⑯　試薬発注支援（ＳＰＤ）システムを準備すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料の提供期限　令和５年12月８日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒560―8552 大阪府豊中市刀根山５－１－１　独立行政法人国立病院機構大阪刀根山医療センター　企画課契約係　山口　柚実　電話06―6853―2001　内線7306
メールアドレス
yamaguchi.yumi.um@mail.hosp.go.jp

3　導入説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和５年10月24日から令和５年12月８日まで。
⑵　交付方法　２⑵に示す場所にて直接交付するか、若しくは電子メールにて交付する。電子メールでの交付を受ける場所は２⑵に示すメールアドレスに調達件名・会社名・担当者名・連絡先を入力し、参加希望の旨メール送信すること。

4　その他　この導入契約の詳細は導入説明書による。なお、本公表内容は予定であり、変更する事があり得る。

5　Summary
⑴　Classification of the products to be procured : 4, 22, 31
⑵　Nature and quantity of the products to be purchased : Composite contract for inspection equipment and testing reagent 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The specimen inspection equipment should have barcode label reading processing function attached to the specimen and it is possible to perform automatic setting based on the order information and bidirectional data communication with the inspection system
②　The specimen inspection equipment should have correspond to items currently being measured regarding biochemistry, immunity, blood, hemostasis, blood sugar, hemoglobin A1c and generally is few change in the reference value, the correlation function is 1.0, the intercept is close to 0.0. In addition, the specimen inspection equipment also corresponds to items currently being measured in the field of general urine, with few changes in measured values
③　There is no change in the inspection material as the model change
④　Be as power-saving and space-saving as possible
⑤　The amount of specimen required for examination is small
⑥　It should be few troubles of equipment, and be possible to recover by prompt response 365 days as necessary, in case trouble occurs in equipment
⑦　A contact point (customer support center, etc.) must be established to respond to questions about the equipment, data problems, academic support, etc.
⑧　Be able to respond when training is necessary for equipment use and provide technical guidance to operators on-site as necessary
⑨　Including the cost of training to provide technical guidance to the equipment managers and operators on the procured items
⑩　Even if it is not specified in the specifications, all matters necessary for the operation of the procured items shall be carried out without fail and if you have any doubts, please consult with the person in charge of our center and receive instructions
⑪　In the event of a failure due to normal use, the equipment must be repaired free of charge for one year after the start of operation
⑫　Consist of equipment and reagents with sufficient consideration for the revised medical therapy (clinical laboratories)
⑬　If the operation method differs from the current equipment, create a manual similar and the like in the current manual, and take care to ensure that there are no problems with daily operations or after-hours operations
⑭　If all of the equipment procured this time requires pure water production equipment, consolidate to two units (equipment from the same manufacturer) with multiple ion exchange resins per unit of equipment and prepare a system that allows the mutual use of pure water for equipment maintenance and in case of trouble
⑮　In principle, uninterruptible equipment should be provided for procured equipment. And the pure water equipment and control terminals are also equipped with an uninterruptible equipment to handle momentary power outages throughout the entire system
⑯　To prepare a reagent ordering support system (SPD)
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M. December 8, 2023
⑹　Contact point for the notice : Yamaguchi Yumi, Contract Clerk, Planning Division, National Hospital Organization Osaka Toneyama Medical Center, 5―1―1 Toneyama, Toyonaka-shi, Osaka-fu, 560―8552, Japan. TEL 06―6853―2001 ext. 7306

## 意見招請に関する公示

次のとおり調達物品等の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和５年10月24日
支出負担行為担当官
国税庁長官官房会計課長　小平　武史
◎調達機関番号　015　◎所在地番号　13
1　調達内容
⑴　品目分類番号　71、27
⑵　購入等特定役務及び数量　令和６年度免税販売管理システムの開発及び改修　一式
2　意見の提出方法
⑴　意見の提出期限　令和５年11月16日17時00分
⑵　提出先　〒100―8978 東京都千代田区霞が関３－１－１　東受付室内入札説明室　（郵送の場合は下記５）
3　仕様書案の交付
⑴　交付期間　令和５年11月16日まで
⑵　交付場所　上記２⑵に同じ
ただし、仕様書案は「政府電子調達システム」を利用して入手することができる。
4　問い合わせ先　国税庁長官官房会計課契約第一係　小林　丈泰　電話03―3581―4161　内線3644