C　手術室内映像コントロールシステムは、システム本体、操作パネル、術野カメラ、ディスプレイ、スイッチャ、入出力パネル等で構成されていること。

D　術野カメラは、４Ｋ・ＦＨＤに対応したズーム・回転・先回等の機能を有するカメラを本院のカメラアームに取り付けること。

E　術野映像記録配信システムは、選択された術野映像を患者情報と紐つけて記録を行い、サーバにて高画質保存管理することができ、端末からは閲覧、ライブ配信、編集、ダウンロード等が行えること。閲覧・ライブ配信については専用端末だけではなく院内のＨＩＳ端末からも可能であること。

F　術野映像記録配信システムは、記録装置、操作パネル、保存管理サーバ、専用端末等で構成されていること。また、院内のＨＩＳネットワークへ接続し、患者情報や手術情報を病院情報システムとデータ連携して取得できること。

G　術場映像記録配信システムは、手術部内の監視カメラ映像や術野映像、生体情報モニタ映像等を、院内各所の端末に配信して分割表示でき、また一定期間保存管理が行えること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年１月12日17時15分（郵送の場合は必着のこと。）

⑵　提供先　〒390―8621 長野県松本市旭３―１―１　信州大学医学部附属病院経営管理課　紙谷　康弘　電話0263―37―2735

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和５年12月１日から令和６年１月12日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和５年12月８日13時30分

⑵　開催場所　信州大学医学部附属病院外来棟研修室７

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 15, 28

⑵　Nature and quantity of the products to be purchased : Surgical Video Recording and Distribution System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The surgical video recording and distribution system shall consist of an operating room video control system, an operating field video recording and distribution system, and an operating field video recording and distribution system.

B　The operating room video control system is a system that can centrally manage the display, recording, distribution, switching, etc. of the surgical field images, patient information monitor images, external input images, etc. in each operating room.

C　The operating room video control system consists of the system itself, the operation panel, the surgical field camera, the display, the switcher, the input／output panel, etc.

D　The surgical field camera should be attached to the camera arm of the main hospital with functions such as zoom, rotation, and forward support for 4K and FHD.

E　The surgical field video recording and distribution system records the selected surgical field video in conjunction with the patient information, and can be stored and managed in high quality on the server, and can be viewed, live-streamed, edited, downloaded, etc. from the terminal. Viewing and live streaming are possible not only from dedicated terminals but also from hospital information system terminals in the hospital.

F　The surgical field video recording and distribution system shall consist of a recording device, an operation panel, a storage management server, a dedicated terminal, and the like. In addition, it is possible to connect to the hospital information system network in the hospital and obtain patient information and surgery information by linking data with the hospital information system.

G　The surgical field video recording and distribution system can distribute surveillance camera images, surgical field images, biometric information monitor images, etc. in the surgical department to terminals in various parts of the hospital for divided display, and can be stored and managed for a certain period of time.

⑸　Time limit for the submission of the requested material : 17 : 15　12 January, 2024

⑹　Contact point for the notice : Yasuhiro Kamiya, Procurement Section, Accounting Division, Shinshu University Hospital, 3―1―1 Asahi Matsumoto-shi Nagano 390―8621 Japan, TEL 0263―37―2735

## 資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和５年12月１日

国立大学法人神戸大学

契約担当役　理事　吉田　　潔

◎調達機関番号　415　◎所在地番号　28

**○第８号**

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　臨床検査分析・支援システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和６年度７月以降

⑸　調達に必要とされる基本的な要求要件

１）検体検査搬送システム

①　上位の検査情報システムと連携できること。

②　検体１本毎に迅速に搬送でき、24時間稼働できること。

③　自動的に遠心分離操作ができること。

④　調達および既存の分析装置に接続でき、迅速かつ効率的に検査が行えること。

⑤　24時間受付部、検体搬入・仕分け部、検体前処理部、生化学・免疫搬送ライン、血液搬送ライン、血糖凝固搬送ラインおよび検体保存部を有すること。

⑥　検体保存部は血清検体を冷蔵で保存できること。

２）臨床検査分析システム

以下の分析装置を含むこと。

①　全自動生化学汎用自動分析装置（電解質分析機能を含む）

②　全自動多項目血球分析システム（自動塗抹標本作製、血液像自動分析装置を含む）

③　全自動血液凝固測定装置

④　グリコヘモグロビン分析装置

⑤　血液ガス分析装置（管理システムを含む）

⑥　全自動浸透圧分析装置

⑦　自動遺伝子解析装置

⑧　遺伝子解析装置

⑨　高圧蒸気滅菌装置

３）採血業務支援システム

既存の自動採血管準備システム、検体・生理検査システムと連携できること。

４）超音波画像診断システム

超音波検査が行え、既存の上位生理検査情報システムと連携できること。

５）検体検査情報処理システム

①　検体検査（細菌検査を含む）の業務が行えること。

②　既存の検査情報システムのデータを移行できること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年１月16日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒650―0017 神戸市中央区楠町７―５―１　神戸大学医学部管理課病院契約係　岸本　知久　電話078―382―5120

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和５年12月１日から令和６年１月16日

⑵　交付場所　上記２⑵に同じ。