4　説明会の開催　本公表に基づく導入説明会を開催する。
(1)　開催日時　令和５年12月15日11時00分
(2)　開催場所　宮崎大学木花キャンパス事務局３階会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1)　Classification of the products to be procured : 22, 31
(2)　Nature and quantity of the products to be purchased : Biochemistry and immunological testing system 1 Set
(3)　Type of the procurement : Purchase
(4)　Basic requirements of the procurement :
A　Biochemistry analyzer-one unit
(a)　Must be capable of performing simultaneous multi-item analysis using a discrete method.
(b)　The instrument should process a maximum of 5,400 tests per hour for the photometric method and maximum of 1,800 tests per hour for the electrolytic method.
B　Specimen Pretreatment Conveyance System-one unit
(a)　Must be able to connect to multiple biochemical autoanalyzers and immunochemiluminescence analyzers.
(b)　The system should be capable of processing more than 500 specimens per hour at a 1 : 1 dispensing ratio.
(c)　The dispensing area should be able to detect liquid levels.
(d)　It should be able to dispense 200—800 μℓ at once.
(e)　The dispensing area of the system should have droplet prevention function.
(f)　The system can store more than 400 processed specimens.
(g)　It should be able to replenish consumables even during operation.
(h)　It should be selectable of the right／left direction where the specimens are conveyed to.
(i)　It should have a backup device.
C　Immunoassay system-one unit
(a)　The system should have a processing capability of 200 tests per hour.
(b)　Sample capacity should be more than 150 samples.
(c)　It should have an automatic re-examination function.
(d)　IT should have a backup device with equal performance to the main unit.
D　Seismic isolator device-one unit
(a)　The device should be installed with seismic isolation.
(b)　The height of the seismic isolator should be less than 10 mm.
(c)　It should withstand a superimposed load of more than 500 kg per square meter.
(d)　It does not need regular maintenance.
E　Temperature Control System-one unit
(a)　It can use cloud servers.
(b)　The operational status of the equipment can be checked in real time.
(c)　Data should be downloadable in CSV format.
F　Benchtop Centrifuge-one unit
(a)　Maximum rotation per minute should be more than 3,500 rpm. The maximum centrifugal force should be more than 2,300×g.
G　Floor-Standing Refrigerated Centrifuge-one set
(a)　Maximum rotation per minute should be more than 4,800 rpm. The maximum centrifugal force should be more than 4,150×g.
(5)　Time limit for the submission of the requested material : 17 : 00　26 January, 2024
(6)　Contact point for the notice : Takanori Kurogi, Procurement Section, Accounting Division, National University Corporation University of Miyazaki, 1—1 Gakuen Kibana-dai Nishi Miyazaki-shi 889—2192 Japan, TEL 0985—58—7574

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和５年12月11日
国立大学法人宮崎大学
契約担当役　理事　佐藤　一仁
◎調達機関番号　415　◎所在地番号　45
○第４号
1　調達内容
(1)　品目分類番号　22、31
(2)　導入計画物品及び数量　Ｘ線ＣＴ組合せ型ＳＰＥＣＴ装置　一式
(3)　調達方法　購入
(4)　導入予定時期　令和６年８月以降
(5)　調達に必要とされる基本的な要求要件
Ａ　ＳＰＥＣＴとＸ線ＣＴを搭載したＳＰＥＣＴ－ＣＴ装置であること。
Ｂ　ＳＰＥＣＴ画像とＣＴ画像の融合を高精度で行う機能を有すること。
Ｃ　ＣＴ画像による減弱補正、散乱補正を行う事により定量値を測定できること。
Ｄ　診断のための各種画像解析及び機能解析ができる機能を有すること。
Ｅ　当院の放射線情報システム・電子カルテシステムと接続し、画像情報・オーダー情報の通信ができること。
Ｆ　迅速なメンテナンス体制とユーザー支援体制がとれていること。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1)　資料等の提供期限　令和６年１月26日17時00分（郵送の場合は必着のこと。）
(2)　提供先　〒889—2192 宮崎市学園木花台西一丁目１番地　国立大学法人宮崎大学財務部経理調達課　黒木　高徳　電話0985—58—7574
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1)　交付期間　令和５年12月11日から令和６年１月26日まで。
(2)　交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1)　開催日時　令和５年12月15日14時00分
(2)　開催場所　宮崎大学木花キャンパス事務局３階会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
(1)　Classification of the products to be procured : 22, 31
(2)　Nature and quantity of the products to be purchased : SPECT-CT system 1 Set
(3)　Type of the procurement : Purchase
(4)　Basic requirements of the procurement :
A　The SPECT-CT equipment must have both SPECT and X-ray CT capabilities.
B　The system must have the ability to fuse SPECT images and CT images with high precision.
C　Quantitative values can be measured by performing attenuation correction and scattering correction using CT images.
D　Must have the ability to perform various image analyzes and functional analyzes for diagnosis.
E　Able to connect with our hospital's radiation information system and electronic chart system and communicate image information and order information.
F　A prompt maintenance system and user support system must be in place.
(5)　Time limit for the submission of the requested material : 17 : 00　26 January, 2024
(6)　Contact point for the notice : Takanori Kurogi, Procurement Section, Accounting Division, National University Corporation University of Miyazaki, 1—1 Gakuen Kibana-dai Nishi Miyazaki-shi 889—2192 Japan, TEL 0985—58—7574

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和５年12月11日
東北管区警察局岩手県情報通信部長
菊地　光一