⑷ Basic requirements of the procurement :

① Multi-slice Computed Tomography System

・The system should be capable of scanning more than 16 slices data during one rotation

・The system should have ability that could scan either whole lung or abdomen in few seconds

② Magnetic Resonance Imaging System

・Magnetic field strength should be 1.5 Tesla or more

・The images of examination in head, body and extremities are enable to examine speedily with high resolution images

・The console unit must provide high performance capability and exceptional in operation

③ Serial Angiographic System

・The system should be designed safely for interventional proce-dures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography

・C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary-direction

・The flat-panel detector should be especially designed to fulfill the interventional angiography

・The digital imaging processing system should have supporting and analysis function for intravascular treatment

④ Gamma Camera System

・The system should be multiple digital gamma camera

・The system should have function of SPECT and Whole body acquisition

・The system should equip the fusion software and quantitative application replete

⑤ High Radiation Therapy System

・The system should be capable to multipurpose treatment

・The system should equip multi-leaf collimator with high shielding performance

・The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiation therapy

⑥ X-ray fluoroscope Radiography System

・The system is available in fluoroscopic radiography

・The system should be capable of recording dynamic images

・The system should be capable of positioning the bed position to horizontal

・The system should equip the flat-panel detector

⑦ X-ray generator System

・The system should equip X-ray tube

・The system should equip a table for standing and lying position, in addition, radiography of whole body could be able to perform speedily and safely

⑧ Flat-panel detector System

・The system should equip the flat-panel detector

⑨ Positoron Emission Tomography-CT

・The PET scanner should be combined with multi-slice X-ray computed tomography

・Image fusion as PET+CT, and PET+MRI should be able to proceed

⑩ Digital X rays equipment to mammography

・The system should be digital X-rays equipment to mammography

・The system must be able to digital stereotactic mammography

・The system should equip the high resolution monitor system for mammography diagnosis

⑪ Mobile X-ray Surgical Imaging System

・The system should be digital mobile X-ray surgical imaging system

⑸ Time-limit for the submission of the requested material : 5 : 00 P.M. March 4, 2024

⑹ Contact point for the notice : Takaki Fukasawa, Chief Clerk, Supply Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2—5—21 Higashigaoka Meguro-ku Tokyo 152—8621, Japan. TEL 03—5712—5065

意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和６年１月18日

独立行政法人大学入試センター

契約担当役　理事　西井　知紀

◎調達機関番号　506　◎所在地番号　13

１　調達内容

⑴　品目分類番号　76

⑵　購入等特定役務及び数量　大学入学共通テスト出願受付等業務　一式

２　意見の提供方法

⑴　意見の提出期限　令和６年２月８日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒153—8501 東京都目黒区駒場２丁目19番23号　独立行政法人大学入試センター総務部財務課契約調達係　篠田　友香　電話03—5478—1225

３　仕様書案の交付

⑴　交付期間　令和６年１月18日から令和６年２月８日まで。

⑵　交付場所　上記２⑵に同じ

４　仕様書案の説明会

⑴　開催日時　令和６年１月26日14時00分

⑵　開催場所　独立行政法人大学入試センター１階会議室

５　Summary

⑴　Classification of the services to be required : 76

⑵　Nature and quantity of the services to be required : Publishing the brochure of National Center Test for University Admissions, accepting application forms, and inputting the data

⑶　Time limit for the submission of comments : 17 : 00　8, February, 2024

⑷　Contact point for the notice : Yuka Shinoda for Financial Affairs Section, Operations Planning Division, The National Center for University Entrance Examinations 2—19—23 Komaba Meguro-ku Tokyo 153—8501 Japan TEL 03—5478—1225

意見招請に関する公示

次のとおり調達案件の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和６年１月18日

契約担当職

国立研究開発法人産業技術総合研究所

大型調達室長　山野　雅史

◎調達機関番号　808　◎所在地番号　08

○産総第 246 号

１　調達内容

⑴　品目分類番号　24

⑵　購入等件名及び数量

高速シリコン除去モジュール　一式

２　意見の提出方法

⑴　意見の提出期限　令和６年２月８日17時（郵送の場合は必着のこと。）

⑵　提出先　〒305—8560 茨城県つくば市梅園１－１－１　国立研究開発法人産業技術総合研究所　大型調達室　藤澤　昌弘　電話029—861—2740

３　仕様書案の交付

⑴　交付期間　令和６年１月18日から令和６年２月２日まで。

⑵　交付場所　上記２⑵に同じ。

４　仕様書案の説明会　実施しない。

５　Summary

⑴　Classification of the products to be procured : 24

⑵　Nature and quantity of the products to be purchased : Silicon rapid-removal module 1 Set

⑶　Time-limit for the submission of the comments : 17 : 00, February 8, 2024

⑷　Contact point for the notice : FUJISAWA Masahiro, High-value Procurement Office, National Institute of Advanced Industrial Science and Technology, 1—1—1 Umezono, Tsukuba, Ibaraki, 305—8560 Japan. TEL 029—861—2740