2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和６年４月26日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学財務部経理課　五十嵐章浩　電話0776―61―3111　内線3145
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和６年３月11日から令和６年４月26日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　開催しない。本件調達に関する問合せは個別に対応する。
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Hybrid OR system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The angiography device shall be a ceiling-suspended single-arm type with a wide range of motion in the vertical and lateral directions.
B　The system must be able to be used by multiple clinical departments.
C　In order to support transcatheter aortic valve implantation and percutaneous left atrial appendage closure, it has a real-time fusion function between fluoroscopic images and ultrasound diagnostic equipment, making it possible to quickly and easily select device size and consider access routes.
D　In order to reduce the radiation exposure of patients and operators, we must have radiation reduction technology whose radiation exposure reduction rate has been proven by clinical data.
E　Equipped with the latest applications, including applications that enhance stent visualization.
F　As an educational tool for young doctors, it is possible to perform simulations using 3D images before surgery, and has a multimodality fusion function that includes marking functions on fluoroscopic and photographic images, and echo images that allow understanding of anatomical structures on fluoroscopy.
G　The system can be used for research and development as a facility that provides highly advanced medical care.
⑸　Time limit for the submission of the requested material : 17 : 00　26 April, 2024
⑹　Contact point for the notice : Igarashi Akihiro, Procurement Section, Accounting Division, Finance Department, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3145

○第６号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　血管造影Ｘ線診断装置　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和７年３月
⑸　調達に必要とされる基本的な要求要件
A　血管撮影装置は天吊り式シングルアームであること。
B　複数の診療科で使用できるようなシステムであること。
C　Cone-Beam CT撮影機能を有すること。
D　リアルタイム３Ｄ再構成機能を有し、迅速かつ簡便にデバイスサイズの選択やアクセスルートの検討が可能であること。
E　患者、術者の被ばくを低減するため、被ばく低減率が臨床データにより立証された被ばく低減技術を有すること。
F　ステント描出を高めるアプリケーションなど最新のアプリケーションを搭載していること。
G　高度先進医療を提供する施設として、研究・開発に利用できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和６年４月26日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学財務部経理課　五十嵐章浩　電話0776―61―3111　内線3145
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和６年３月11日から令和６年４月26日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　開催しない。本件調達に関する問合せは個別に対応する。
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Angiography X-ray diagnostic equipment 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The angiography device must be a ceiling-suspended single arm.
B　The system must be able to be used by multiple clinical departments.
C　Must have Cone-Beam CT imaging function.
D　It has a real-time 3D reconstruction function, and it is possible to quickly and easily select device size and consider access routes.
E　In order to reduce the radiation exposure of patients and operators, we must have radiation reduction technology whose radiation exposure reduction rate has been proven by clinical data.
F　Equipped with the latest applications, including applications that enhance stent visualization.
G　The facility can be used for research and development as a facility that provides highly advanced medical care.
⑸　Time limit for the submission of the requested material : 17 : 00　26 April, 2024
⑹　Contact point for the notice : Igarashi Akihiro, Procurement Section, Accounting Division, Finance Department, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3145

○第７号

1　調達内容
⑴　品目分類番号　14
⑵　導入計画物品及び数量　総合医療情報システム　一式
⑶　調達方法　借入
⑷　導入予定時期　令和８年３月
⑸　調達に必要とされる基本的な要求要件
A　全体のシステム構成として、本院の診療業務を支援し、病院業務を円滑に運営・管理できるとともに、全体の情報セキュリティ管理機能を備えた構成であること。
B　診療に関する機能として、患者基本情報、病名歴、初診・経過記録などに関する電子カルテ機能、診療予約、各種検査予約、検査、処方、注射、処置、輸血、手術、看護、食事などに関する指示及びオーダーエントリー機能、各種診療行為に関する実施記録機能を含んでいること。
C　部門サブシステムの機能として、会計、薬剤、検査、放射線、輸血、手術、栄養給食、医療安全、感染制御、リハビリテーション、地域連携などの病院内部門の業務を取り扱えること。また、各システムに登録された各種情報は、診療、診療報酬請求、経営など病院全体でデータが活用できる機能を備えること。
D　電子的記録については、個人情報保護に資する情報システムの運用管理への配慮やｅ‐文書法等への適切な対応として「医療情報システムの安全管理に関するガイドライン第６版」（以下「ガイドライン」という。）を遵守した運用及び説明が可能な法的にも安全なシステム構成であること。