C　The mirror shall have a mechanism that allows a 4K3D camera system to be built into the base of the mirror.

D　The auto-balance adjustment should be able to adjust the balance no matter which position the mirror base is placed.

E　The stand system shall have interface terminals for navigation.

⑸　Time limit for the submission of the requested material : 17 : 00　3 June, 2024

⑹　Contact point for the notice : Tetsuo Kikuchi, Administration Section, Administration Division, Tokyo Medical and Dental University Hospital, 1―5―45 Yushima Bunkyo-ku Tokyo 113―8519 Japan, TEL 03―5803―5410

○第３号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　人工心肺装置　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和６年度１月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　全てのポンプがバッテリーにて駆動できること。

Ｂ　安全機能である、圧力モニタ、貯血槽のレベル検知機能、気泡検知機能を有すること。

Ｃ　心筋保護装置機能を有すること。

Ｄ　日本体外循環医技術医学会（JaSECT）から発行されている、「人工心肺における安全装置設置基準」を遵守した機能を有すること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年６月３日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒113―8519 東京都文京区湯島１―５―45　東京医科歯科大学病院管理課新棟・再整備準備室調達・移転実施係　菊地　哲生　電話03―5803―5410

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和６年４月16日から令和６年６月３日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和６年４月23日15時00分

⑵　開催場所　東京医科歯科大学病院管理課打合せ室（M＆Dタワー１階）

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Heart-Lung Machine 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　All pumps shall be able to be driven by battery alternatively.

B　Safety features such as pressure monitoring, reservoir level detection function, and bubble detection function shall be provided.

C　The device shall have the cardioplegia circulation function.

D　The device shall have safety functions which comply with “Installation Standards of Safety Devices during Cardiopulmonary Bypass” issued by The Japanese Society of Extra- Corporeal Technology in Medicine.

⑸　Time limit for the submission of the requested material : 17 : 00　3 June, 2024

⑹　Contact point for the notice : Tetsuo Kikuchi, Administration Section, Administration Division, Tokyo Medical and Dental University Hospital, 1―5―45 Yushima Bunkyo-ku Tokyo 113―8519 Japan, TEL 03―5803―5410

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和６年４月16日

国立大学法人福井大学財務部長　堂　　裕司

◎調達機関番号　415　◎所在地番号　18

○第２号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ＦＰＤ搭載Ｘ線撮影システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和７年３月

⑸　調達に必要とされる基本的な要求要件

Ａ　天井走行式Ｘ線管保持装置を有し、立位用Ｘ線撮影装置、臥位用Ｘ線撮影装置を備えること。

Ｂ　パワーアシスト機能を有すること。

Ｃ　間接方式（蛍光体方式）の平面検出器が搭載されていること。

Ｄ　ワイヤレスフラットパネルディテクタが搭載可能であること。

Ｅ　患者被ばく低減や被ばく管理を考慮した装置であること。

Ｆ　既存のＲＩＳ・ＰＡＣＳ等の画像システムと接続すること。

Ｇ　読影支援を有すること。

Ｈ　高度先進医療を提供する施設として、研究・開発に利用できること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年５月31日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学財務部経理課　飯田　沙織　電話0776―61―3111　内線3171

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和６年４月16日から令和６年５月31日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　開催しない。本件調達に関する問合せは個別に対応する。

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : General radiographic X-ray apparatus equipped with flat panel detector (FPD) 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　It shall have a ceiling-running X-ray tube holding device, and be equipped with an X-ray imaging device for standing position and an X-ray imaging device for supine position.

B　Must have power assist system.

C　Equipped with an indirect type (fluorescent type) flat panel detector.

D　Must be able to mount a wireless flat panel detector.

E　The device must be designed with consideration for patient exposure reduction and exposure management.

F　Connect with existing image systems such as RIS and PACS.

G　Must have radiology report supporting system.

H　The facility can be used for research and development as the medical university that provides highly advanced medical care.

⑸　Time limit for the submission of the requested material : 17 : 00　31 May, 2024

⑹　Contact point for the notice : Iida Saori, Procurement Section, Accounting Division, Finance Department, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3171