The X-rays detector mechanism being a flat panel method.

Have interface for DICOM.

Be linked by the hospital information system and data communication.

For catheter operation, have support systems such as indication or the image processing of other modality images.

⑸　Time limit for the submission of the requested material : 17 : 00　17 June, 2024

⑹　Contact point for the notice : Kishimoto Tomohisa, Contract Affairs Section, Financial Management Division, Kobe University Hospital, 7―5―1 Kusunoki-cho Chuo-ku Kobe-shi 650―0017 Japan, TEL 078―382―5120

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和６年５月１日

国立大学法人神戸大学

契約担当役　理事　森山　　睦

◎調達機関番号　415　◎所在地番号　28

○第３号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　集中治療・手術モニタリングシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和６年度12月以降

⑸　調達に必要とされる基本的な要求要件

①　心電図、呼吸曲線、観血血圧波形､脈波、CO_2分圧曲線、脳波、呼吸流量曲線、気道内圧曲線、換気量曲線、脳波（ＢＩＳ）、O_2濃度曲線、CO_2濃度曲線、Agent濃度曲線の波形表示が可能であること

②　心拍数、ＶＰＣ数、ＳＴレベル､呼吸数、非観血血圧値、観血血圧値、SpO２値、脈拍数、体温、心拍出量、心係数、注入液温度、血液温度、呼気終末CO_2分圧、ＢＩＳ値、QTc、QRSdの数値表示が可能であること

③　重症記録システムへの連動が可能であること

④　モニタリングしたパラメーターはサーバー等に記録され、過去の記録を閲覧することが可能であること

⑤　rSO２モジュールの接続が可能であること

⑥　現行接続している医療機器の接続が可能であること

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年６月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒650―0017 神戸市中央区楠町７－５－１　神戸大学医学部管理課病院契約係　岸本　知久　電話078―382―5120

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和６年５月１日から令和６年６月17日

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和６年５月８日14時00分

⑵　開催場所　神戸大学医学部管理棟入札室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Intensive care and operation monitoring system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

①　Availability of displaying ECG, respiratory curve, blood pressure waveform, pulse wave, CO_2 partial pressure curve, EEG, respiratory flow curve, airway pressure curve, ventilation volume curve, EEG (BIS), O_2 concentration curve, CO_2 concentration curve, and Agent concentration curve waveforms

②　Availability of displaying heart rate, VPC count, ST level, respiratory rate, unobserved blood pressure value, observed blood pressure value, SpO2 value, pulse rate, body temperature, cardiac output, cardiac index, infusion fluid temperature, blood temperature, end-expiratory CO_2 partial pressure, BIS value, QTc, QRSd

③　Availability of linkage to the critical illness record system

④　Monitored parameters must be recorded on a server, and past records must be able to be viewed

⑤　The rSO2 module must be able to be connected

⑥　Availability to connect medical devices currently connected to the system

⑸　Time limit for the submission of the requested material : 17 : 00　17 June, 2024

⑹　Contact point for the notice : Kishimoto Tomohisa, Contract Affairs Section, Financial Management Division, Kobe University Hospital, 7―5―1 Kusunoki-cho Chuo-ku Kobe-shi 650―0017 Japan, TEL 078―382―5120

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和６年５月１日

国立大学法人神戸大学

契約担当役　理事　森山　　睦

◎調達機関番号　415　◎所在地番号　28

○第４号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　放射線治療計画システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和６年度11月以降

⑸　調達に必要とされる基本的な要求要件

多様な症例の治療計画画像を取得するために、大口径ボアのＣＴ装置であること。

放射線治療装置と同じ体位が再現できるようにフラット天板、レーザーポインター及び固定具を有すること。

各放射線治療装置の呼吸管理システムに対応した画像が取得できること。

治療計画支援システムを有すること。

高精度治療を支援する検証ツールを有すること。

高精度放射線治療の多様なニーズに対応するため、治療計画とその検証システム、精度管理が滞りなく円滑に実施できること。

放射線治療情報の管理、及び院内院外における医療連携が可能であること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和６年６月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒650―0017 神戸市中央区楠町７－５－１　神戸大学医学部管理課病院契約係　岸本　知久　電話078―382―5120

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和６年５月１日から令和６年６月17日

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和６年５月９日14時00分

⑵　開催場所　神戸大学医学部管理棟入札室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Radiotherapy Treatment Planning system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

Treatment to the large number of cases to be possible, set up high precision, the external radiation therapy device

It should have a flat tabletop, laser pointers, and fixtures to replicate the same position as the radiotherapy device.

It should be able to acquire images compatible with the respiratory management system of each radiotherapy device.