K　分析による感度低下を低減させるためイオン導入部キャピラリはガラス製であること。

L　温度変化から生じる質量誤差を抑えるため、チラーを用いて制御する機構を有すること。

M　装置の維持及び稼働に液体ヘリウムを要しないこと。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和７年１月27日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒431―3192 浜松市中央区半田山１―20―１　浜松医科大学会計課　和田　哲明　電話053―435―2553

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和６年12月26日から令和７年１月27日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和７年１月８日16時00分

⑵　開催場所　浜松医科大学管理棟２階第二会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 24

⑵　Nature and quantity of the products to be purchased : High-Resolution Mass Spectrometer 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The analyzer must adopt a quadrupole time-of-flight (QTOF) system equipped with ion mobility separation mechanisms.

B　The ionization system must include Electrospray Ionization (ESI), and to allow measurement of unstable samples, the spray needle should not be charged with voltage.

C　For improved maintenance, the ESI system must employ an orthogonal configuration.

D　The system must support Matrix-Assisted Laser Desorption／Ionization (MALDI) and allow analysis using both ESI and MALDI with software-based switching.

E　To acquire high-speed, high-quality, and highly sensitive data, the laser used for MALDI must be a solid-state laser adjustable in the range of 1 to 10,000 Hz.

F　A post-ionization mechanism utilizing a two-stage laser irradiation process must be provided.

G　The ion mobility separation mechanism must calculate Collision Cross-Section (CCS) values.

H　The system must enable imaging mass spectrometry analysis through ion mobility separation, allowing the separation of isomers from each other and target peaks from noise peaks.

I　The mass accuracy of the system must be within 0.8 ppm (with internal standards) and 2 ppm (with external standards).

J　The mass resolving power must be 60,000 or greater (measured by full width at half maximum).

K　To reduce sensitivity loss during analysis, the capillary in the ion introduction section must be made of glass.

L　To minimize mass errors caused by temperature fluctuations, the system must have a mechanism that uses a chiller for control.

M　The system must not require liquid helium for maintenance or operation.

⑸　Time limit for the submission of the requested material : 17 : 00　27 January, 2025

⑹　Contact point for the notice : WADA Tetsuaki, Accounting Division, Hamamatsu University School of Medicine, 1―20―1 Handayama Chuo-ku Hamamatsu city 431―3192 Japan, TEL 053―435―2553

意見招請に関する公示

次のとおり調達物品等の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和６年12月26日

支出負担行為担当官

デジタル庁会計担当参事官　杉本　敬次

◎調達機関番号　027　◎所在地番号　13

1　調達内容

⑴　品目分類番号　71、27

⑵　件名及び数量　電子契約システム（工事・業務）のガバメントクラウド移行に係る設計開発及び結合テスト等業務

2　意見の提出方法

⑴　意見の提出期限　令和７年１月22日10時00分

⑵　提出先　〒102―0094 東京都千代田区紀尾井町１番３号東京ガーデンテラス紀尾井町19階　デジタル庁国民向けサービスグループフロントサービス担当

電子メール：shotyamaz@digital.go.jp

3　仕様書案の交付

⑴　交付期間　令和６年12月26日から令和７年１月22日まで

⑵　交付場所　政府電子調達システムに掲載する。

4　Summary

⑴　Classification of the services to be procured : 71, 27

⑵　Nature and quantity of the Services to be required : Design and development, integration testing, etc. related to the migration of electronic contract systems (construction and business) to the government cloud

⑶　Time-limit for the submission of comments : 10 : 00 AM, 22 January, 2025

⑷　Contact point for the notice : Digital Agency, 1―3 Kioicho, Chiyoda-ku, Tokyo 102―0094 Japan

E-mail : shotyamaz@digital.go.jp

意見招請に関する公示

次のとおり調達物品等の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和６年12月26日

支出負担行為担当官

国税庁長官官房会計課長　小平　武史

◎調達機関番号　015　◎所在地番号　13

1　調達内容

⑴　品目分類番号　71、27

⑵　購入等特定役務及び数量　令和７年度マイナポータル等連携プラットフォームへ追加するアプリケーションの開発等　一式

2　意見の提出方法

⑴　意見の提出期限　令和７年１月16日17時00分

⑵　提出先　〒100―8978 東京都千代田区霞が関３－１－１　東受付室内入札説明室（郵送の場合は下記５）

3　仕様書案の交付

⑴　交付期間　令和７年１月16日まで

⑵　交付場所　上記２⑵に同じ

ただし、仕様書案は「調達ポータル」を利用して入手することができる。

4　仕様書案の説明会

⑴　開催日時　令和７年１月８日13時30分

⑵　開催場所　国税庁第五会議室

5　問い合わせ先　国税庁長官官房会計課契約第一係　谷本　和彦　電話03―3581―4161　内線3644

6　Summary

⑴　Classification of the products to be procured : 71, 27

⑵　Nature and quantity of the products to be required : Development for application of Mynaportal platform System 1 set

⑶　Time-limit for the submission of comments : 17 : 00　16 January 2025.

⑷　Contact point for the notice : TANIMOTO Kazuhiko Procurement Section, Accounts Division, Commissioner's Secretariat, National Tax Agency, 1―1 Kasumigaseki 3-chome, Chiyoda-ku, Tokyo 100―8978 Japan. TEL 03―3581―4161 ex. 3644