D　位置合わせ用のレーザーマーカーを備えること
E　病院が保有するシステムと接続できること
④　全身用Ｘ線ＣＴ診断装置
A　Ｘ線管球と検出器が一体となって連続回転する第３世代方式であること
B　Ｘ線検出器列数が実装で128列以上であること
C　心電同期スキャンが可能であること
D　金属アーチファクト低減が可能であること
E　病院が保有するシステムと接続できること

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑹の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和７年３月14日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒114―8564 東京都北区十条台１－５－70　陸上自衛隊補給統制本部衛生部補給計画課補給整備班補給整備班長　電話03―3908―5121　内線4721
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和７年１月27日から令和７年３月13日まで。
⑵　交付場所　上記２⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased :
①　Cardiovascular Interventional Digital Angiography System, stationary type 1 Set
②　Endoscopic surgery Support Device 1 Set
③　X-ray CT for radiation therapy planning 1 Set
④　Whole-body X-ray CT Imaging Device 1 Set
⑶　Type of the procurement : purchase
⑷　Basic performance to be required :
①　Cardiovascular Interventional Digital Angiography System, stationary type
A　It shall have large field flat panel detector (FPD).
B　It shall have a 3D digital subtraction angiography (rotational DSA) function.
C　It must have cone-beam CT function.
D　CT gantry must be sliding type.
E　It shall be possible to be connected to hospital-owned systems.
②　Endoscopic surgery Support Device
A　The system must be medical device to assist endoscopic surgery.
B　It must be remote-controlled by operator.
C　It shall have high-quality enlarged stereoscopic vision, highly flexible operability with multi-joint forceps, and hand-tremor correction function.
D　It must be compatible with robot-assisted surgeries in the field of urology, gastrointestinal surgery, gynecology, and thoracic surgery excluding cardiac surgery, which are covered by insurance medical treatment.
③　X-ray CT for radiation therapy planning
A　An X-ray CT scan system for treatment planning must be included.
B　It must be equipped with scanning detector more than 64 rows.
C　It must be capable of 4D imaging with respiratory gating and with ECG gating.
D　It shall have laser marker for alignment.
E　It should be possible to be connected to the hospital-owned medical information system.
④　Whole-body X-ray CT Imaging Device
A　It must be third generation system that continuously rotates the X-ray tube and detector as one unit.
B　It must be equipped with scanning detector more than 128 rows.
C　It must be capable of ECG-gated scanning.
D　It shall be possible to reduce metal artifact.
E　It should be possible to be connected to the hospital-owned medical information system.
⑸　Time limit for the submission of the requested material : 17 : 00　14 March 2025.
⑹　Contact point for the notice : Section leader, Supply and Maintenance Section, Medical Department, Ground Materiel Control Command, Japan Ground Self-Defense Force, 1―5―70 Jujodai, Kita-ku Tokyo 114―8564 Japan. TEL 03―3908―5121 Ext. 4721

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和７年１月27日
国立大学法人新潟大学長　牛木　辰男
◎調達機関番号　415　◎所在地番号　15
○第４号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　鏡視下手術総合システム　一式
⑶　調達方法　借入
⑷　導入予定時期　令和７年度９月以降
⑸　調達に必要とされる基本的な要求要件
A　先端が４方向に湾曲するフレキシブルスコープにより３Ｄ内視鏡観察が可能なシステムであること。
B　４Ｋ／３Ｄ高画質モニターにより手術画像を４Ｋ高画質で映し出せること。
C　３Ｄシステムおよび４Ｋ高画質システムは、ビデオスコープのみならず、カメラヘッドや軟性鏡も使用可能であること。
D　ＩＲ術中蛍光観察機能を備え、ＩＲ画像は単色光モードとカラーモードを有すること。
E　消化器内視鏡機器により術中に消化管内の観察を可能とすること。
F　内視鏡システム設定や機器操作はタッチパネルモニターで一括操作できること。
G　導入システムにおいて故障発生時の代替品用意に加え、修理保証、保守定期点検が組み込まれていること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和７年３月14日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課専門職員　伊藤　憲興　電話025―227―0972
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和７年１月27日から令和７年３月14日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和７年２月３日15時00分
⑵　開催場所　新潟大学医歯学総合病院西診療棟３階　第３会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Integrated endoscopic surgical system 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requ of the procurement :
A　The flexible scope whose tip is curved in four directions needs to be compatible with 3D endoscopic observation.