下)、撮影範囲（直径５～85mm以上)、最大撮影可能領域（240mm以上)、最大サンプルサイズ（160φ×400mm以上)、ＣＴ画像画素サイズ（2.9μm／pixel以下)、画像空間分解能（５μm)、サンプルベッド（摘出組織用・マウス用・ラット用・ウサギ用、各２個以上）を備えていること。

ｂ．透視撮影モードから心拍呼吸同期撮影機能を有すること。

ｃ．３Ｄマルチモダリティキットを介してIn Vivo ２Ｄ／３Ｄ 発光・蛍光イメージングシステム（Ａ）で取得した画像データと重ね合わせできること。

Ｃ．ＣＴデータ解析用システム

ａ．２Ｄ及び３ＤのViewerソフトウェア、組織分離機能、３Ｄ体積計測機能を有すること。

ｂ．２種類以上のＢＭＤファントムを有すること。

Ｄ．３Ｄマルチモダリティキット

ａ．In Vivo ２Ｄ／３Ｄ 発光・蛍光イメージングシステム（Ａ）で取得した画像データと高分解能マイクロＸ線ＣＴシステム（Ｂ）で取得した画像データを重ね合わせできること。

ｂ．麻酔下の動物を固定した状態で装置間を移動するための専用カセットを有すること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和７年４月22日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒889―2192 宮崎市学園木花台西一丁目１番地　国立大学法人宮崎大学財務部経理調達課　小田　紗菜　電話0985―58―7574

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和７年３月21日から令和７年４月22日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和７年３月27日14時00分

⑵　開催場所　宮崎大学木花キャンパス事務局３階会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 24

⑵　Nature and quantity of the products to be purchased : In Vivo 2D／3D Bioluminescence, Fluorescence and Micro X-ray CT Multimodality Imaging System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

Ａ．In Vivo 2D／3D Bioluminescence and Fluorescence Imaging System

ａ．A high sensitivity CCD camera system (cooling at below －85℃), excitation filters (10 or more filters), emission filters (18 or more filters), and a scanning laser (output of 1mV or more), capable of acquiring and reconstructing 2D／3D bioluminescence and fluorescence images.

ｂ．The eXcelon-coated back-illuminated detector is quantum efficiency of 90% or more in the wavelength range of 500―70nm and maximum quantum efficiency of 95% or more.

ｃ．Capable of automatically overlaying image data obtained from a high-resolution micro X-ray CT system (B) via the 3D multimodality kit (D).

Ｂ．High resolution Micro X-ray CT System

ａ．Equipped with a tube voltage (maximum of 100kV or more), tube current (maximum of 200μA or more), X-ray focal spot size (5μm or less), a semiconductor flat-panel detector, self-shielding X-ray (leakage dose less than 1μSV／h), imaging range (diameter of 5―85mm or more), maximum imaging area (240mm or more), maximum sample size (160φ×400mm or more), CT image pixel size (2.9μm／pixel or less), image spatial resolution (5μm), and sample beds (2 or more each for excised tissue, mouse, rat, and rabbit).

ｂ．Capable of cardiac and respiratory synchronized imaging from fluoroscopy mode.

ｃ．Capable of overlaying image data obtained from the In Vivo 2D／3D bioluminescence and fluorescence Imaging System (A) via the 3D multimodality kit.

Ｃ．Data analysis system for Micro X-ray CT System

ａ．Equipped with 2D and 3D Viewer software, tissue segmentation, and 3D volume measurement function.

ｂ．Equipped with 2 or more size of BMD phantoms.

Ｄ．3D Multimodality kit

ａ．Capable of overlaying image data obtained from the In Vivo 2D／3D bioluminescence and fluorescence Imaging System (A) and high-resolution micro X-ray CT System (B).

ｂ．Equipped with a dedicated cassette for moving animals under anesthesia between devices.

⑸　Time limit for the submission of the requested material : 17 : 00　22 April 2025

⑹　Contact point for the notice : Suzuna Oda, Procurement Section, Accounting Division, National University Corporation University of Miyazaki, 1―1 Gakuen Kibana-dai Nishi Miyazaki-shi 889―2192 Japan, TEL 0985―58―7574

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和７年３月21日

支出負担行為担当官

内閣府大臣官房会計担当参事官　北村　　実

◎調達機関番号　007　◎所在地番号　13

１　調達内容

⑴　品目分類番号　26

⑵　購入等件名及び数量　全国８か所の備蓄拠点における災害備蓄用調理資機材の製造又は調達及び納品業務

２　意見の提出方法

⑴　意見の提出期限　令和７年４月11日11時00分（郵送の場合は必着のこと。）

⑵　提出先　〒100―8914 東京都千代田区永田町１―６―１　内閣府政策統括官(防災担当)付参事官付（防災デジタル・物資支援担当）　電話03―3503―2231（直通）

３　仕様書案の交付

⑴　交付期間　令和７年３月21日から令和７年４月11日まで。

⑵　交付場所　政府電子調達システムに掲載する。

４　Summary

⑴　Classification of the services to be procured : 26

⑵　Nature and quantity of the services to be required : The manufacturing or procurement and delivery of cooking equipment for disaster stockpiles at eight locations nationwide

⑶　Time-limit for the submission of comments : 11 : 00 AM, 11 April 2025

⑷　Contact point for the notice : TANAKA Megu, Contract 1 Section, Accounts Division, Minister's Secretariat, Cabinet Office, 1―6―1 Nagata-cho, Chiyoda-ku, Tokyo 100―8914 Japan TEL 03―5253―2111 ext. 82356

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和７年３月21日

門司税関総務部長　川﨑　達也

◎調達機関番号　015　◎所在地番号　40

１　調達内容

⑴　品目分類番号　16

⑵　購入等物品及び数量

コンテナ貨物大型Ｘ線検査装置　一式